



T天康医疗防护用品
Tiankang Medical Protective Product



安徽天康(集团)股份有限公司
安徽天康医疗科技股份有限公司

CATALOGUE

目录 ▼

企业概况

Profile of the Company

01/04

资质认证

Certifications

05/40

检验报告

Test Report

41/118

产品展示

Introduction of Products

119/152

生产设备

Introduction of Equipment

153/158



天康医疗防护用品制造部简介:

安徽天康医疗科技股份有限公司（股票简称：天康医疗，股票代码：835942）系国内大型企业集团—安徽天康（集团）股份有限公司于1999年发起创立的下属子公司，专业从事注射器、输液器等100多个品种的医疗耗材产品的研发和生产，并获得CE产品认证、ISO13485国际质量管理体系认证、WHO-PQS认证和美国FDA 510(K)注册等。

新冠肺炎疫情发生以来，在国内医疗防护物资十分紧缺的情况下，公司在集团领导的大力支持下，发挥自己的产业优势于2020年2月份建立防护服制作车间和医用口罩制作车间，专业从事医用一次性防护服、隔离衣、一次性使用医用口罩、医用外科口罩、医用防护口罩、医用隔离眼罩、医用隔离面罩和口腔粘膜消毒液的研发和生产，并获得国内医疗器械注册证、CE产品认证和美国FDA注册列名。目前一次性医用口罩日产量已超1000万只，医用一次性防护服超20万件。

天下人的健康，天康人的追求。欢迎国内外有识之士携手共进，共同发展！

Profile of Tiankang Medical Protective Product Manufacturing Department:



Anhui Tiankang Medical Technology Co., Ltd. (Stock Abbreviation: Tiankang Medical, Stock Code:835942) which was founded as a subsidiary company of Anhui Tiankang (Group) Shares Co., Ltd. in 1999. We specialized in developing and producing more than 100 varieties of medical consumables products such as syringes and infusion sets etc. We have got CE certificate, ISO13485 certificate, WHO-PQS certificate and US FDA 510(K) registration etc.

Since the outbreak of the NCP epidemic, in the case of a shortage of domestic medical protective products, with the strong support of the group leaders, our company has taken our own industry advantages and established protective clothing production workshop and medical mask production workshop in February 2020, we specialized in developing and producing medical coveralls, non-woven isolation gowns, single-use medical face mask, surgical mask, protective face mask for medical use, medical isolation goggles, medical isolation face-shield and oral mucosa disinfectant. We have got domestic medical device registration certificates, CE certificate and US FDA listing number. At present, the daily output of single-use medical face mask has exceeded 10 million pieces, and the daily output of medical coveralls has exceeded 200,000 pieces.

Tiankang is always working for the health of all the people. Warmly welcome men of insight home and abroad to join us and we shall go forward together!





营业执照
Business License



医疗器械生产许可证
Manufacturing License



商标证
Trademark Certificate



医疗器械生产产品登记表

Registration Form for Production of Medical Device

医疗器械生产产品登记表



| | |
|---------|---|
| 企业名称 | 安徽天康医疗科技股份有限公司 |
| 许可证编号 | 皖食药监械生产许 20150052 号 |
| 许可证有效期限 | 2021 年 6 月 12 日 |
| 生产范围 | III类：6866 医用高分子材料及制品 6845 体外循环及血液处理设备 III类、II类：6815 注射穿刺器械 II类：14-13 手术室感染控制用品 14-14 医护人员防护用品 17-04 口腔治疗器具 |

生产产品列表

| 序号 | 产品名称 | 注册号 | 登载日期 | 备注 |
|----|-----------------------|------------------|------------|----|
| 1 | 一次性使用输液器带针 | 国械注准 20153661941 | 2020.03.10 | |
| 2 | 一次性使用输血器 | 国械注准 20183661721 | 2020.03.10 | |
| 3 | 一次性使用无菌注射器带针 | 国械注准 20143152353 | 2020.03.10 | |
| 4 | 一次性使用无菌注射针 | 国械注准 20193141998 | 2020.03.10 | |
| 5 | 一次性使用静脉留置针 | 国械注准 20173151032 | 2020.03.10 | |
| 6 | 一次性使用无菌自毁型固定剂量疫苗注射器带针 | 国械注准 20173150989 | 2020.03.10 | |
| 7 | 一次性使用无菌自毁式注射器带针 | 国械注准 20173150630 | 2020.03.10 | |
| 8 | 一次性使用无菌回缩型自毁式注射器带针 | 国械注准 20173150631 | 2020.03.10 | |

| 序号 | 产品名称 | 注册号 | 登载日期 | 备注 |
|----|-----------------|------------------|------------|----|
| 9 | 一次性使用溶药器带针 | 皖械注准 20182150065 | 2020.03.10 | |
| 10 | 一次性使用精密过滤输液器带针 | 国械注准 20173664007 | 2020.03.10 | |
| 11 | 一次性使用无菌胰岛素注射器 | 国械注准 20173151045 | 2020.03.10 | |
| 12 | 一次性使用静脉输液针 | 国械注准 20163151175 | 2020.03.10 | |
| 13 | 一次性使用无菌自毁式注射器带针 | 国械注准 20193141872 | 2020.03.10 | |
| 14 | 一次性使用动静脉瘘穿刺针 | 国械注准 20153152195 | 2020.03.10 | |
| 15 | 血液净化装置的体外循环血路 | 国械注准 20153452194 | 2020.03.10 | |
| 16 | 一次性使用静脉输液针 | 国械注准 20163151175 | 2020.03.10 | |
| 17 | 口腔种植一次性使用输水管 | 皖械注准 20192170111 | 2020.03.10 | |
| 18 | 一次性使用医用口罩 | 皖械注准 20202140058 | 2020.03.10 | |
| 19 | 医用外科口罩 | 皖械注准 20202140059 | 2020.03.10 | |
| 20 | 医用一次性防护服 | 皖械注准 20202140060 | 2020.03.10 | |

备注：一次性使用医用口罩、医用外科口罩、医用一次性防护服为新型冠状病毒肺炎应急防护用品，注册证有效期一年，应按照产品注册证要求，加强上市后产品原材料供应商审计和产品质量控制。

发证部门：安徽省药品监督管理局



医疗器械注册证 (一次性使用医用口罩)

Registration Certificates (Single-use Medical Face Mask)

医疗器械注册证 (医用外科口罩)

Registration Certificates (Surgical Mask)

中华人民共和国医疗器械注册证

注册证编号: 皖械注准 20202140058

| | |
|-------|--|
| 注册人名称 | 安徽天康医疗科技股份有限公司 |
| 注册人住所 | 安徽省天长市经济开发区纬一路 228 号 |
| 生产地址 | 安徽省天长市经济开发区纬一路 228 号 |
| 产品名称 | 一次性使用医用口罩 |
| 型号、规格 | 按口罩带的型式分为系带式 and 松紧带式两种, 按灭菌方式分为灭菌型 and 非灭菌型两种型号; 按尺寸分为: 17cm×8cm、17cm×9cm、17cm×9.5cm、17cm×10cm、17.5cm×8cm、17.5cm×9cm、17.5cm×9.5cm、17.5cm×10cm、18cm×8cm、18cm×9cm、18cm×9.5cm、18cm×10cm、19.5cm×8cm、19.5cm×9cm、19.5cm×9.5cm、19.5cm×10cm、12.5cm×9cm、14.5cm×9cm、16.5cm×9cm 十九个规格。 |
| 结构及组成 | 由口罩体、鼻夹及口罩带组成。灭菌型产品经环氧乙烷灭菌, 产品应无菌。 |
| 适用范围 | 用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。 |
| 附件 | / |
| 其他内容 | / |
| 备注 | 1. 本产品为新型冠状病毒肺炎应急防护用品, 注册证有效期一年。 2. 注册人需在延续注册时按照补正通知要求完善生物相容、产品有效期和包装研究资料。 3. 上市后加强产品原材料供应商管理和产品质量控制。 |

审批部门: 安徽省药品监督管理局

批准日期: 2020年03月05日
有效期至: 2021年03月04日



中华人民共和国医疗器械注册证

注册证编号: 皖械注准 20202140059

| | |
|-------|---|
| 注册人名称 | 安徽天康医疗科技股份有限公司 |
| 注册人住所 | 安徽省天长市经济开发区纬一路 228 号 |
| 生产地址 | 安徽省天长市经济开发区纬一路 228 号 |
| 产品名称 | 医用外科口罩 |
| 型号、规格 | 按口罩带的型式分为系带式和耳挂式两种, 按灭菌方式分为灭菌型 and 非灭菌型两种型号; 按尺寸分为: 12.5cm×9cm、14.5cm×9cm、16.5cm×9cm、17cm×8cm、17cm×9cm、17cm×9.5cm、17.5cm×8cm、17.5cm×9cm、17.5cm×9.5cm、17.5cm×10cm、18cm×8cm、18cm×9cm、18cm×9.5cm、18cm×10cm、19.5cm×8cm、19.5cm×9cm、19.5cm×9.5cm、19.5cm×10cm 十九个规格。 |
| 结构及组成 | 由无纺布层、熔喷布、鼻夹与口罩带组成。灭菌型产品经环氧乙烷灭菌, 产品应无菌。 |
| 适用范围 | 供临床医务人员在有创操作过程中佩戴, 覆盖住使用者的口、鼻及下颌, 为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。非灭菌型产品应灭菌后使用。 |
| 附件 | / |
| 其他内容 | / |
| 备注 | 1. 本产品为新型冠状病毒肺炎应急防护用品, 注册证有效期一年。 2. 注册人需在延续注册时按照补正通知要求完善生物相容、产品有效期和包装研究资料。 3. 上市后加强产品原材料供应商管理和产品质量控制。 |

审批部门: 安徽省药品监督管理局

批准日期: 2020年03月05日
有效期至: 2021年03月04日



医疗器械注册证（医用防护口罩）

Registration Certificates (Protective Face Mask for Medical Use)

中华人民共和国医疗器械注册证

注册证编号：皖械注准 20202140350

| | |
|-------|--|
| 注册人名称 | 安徽天康医疗科技股份有限公司 |
| 注册人住所 | 安徽省天长市经济开发区纬一路 228 号 |
| 生产地址 | 安徽省天长市经济开发区纬一路 228 号 |
| 产品名称 | 医用防护口罩 |
| 型号、规格 | 折叠型 15.3cm × 10.3cm 按灭菌方式分为灭菌型和非灭菌型两种型号 |
| 结构及组成 | 由口罩体、鼻夹、口罩带、调节扣及密封条组成。灭菌型应无菌。 |
| 适用范围 | 用于戴在医疗机构与病毒物料接触的人员面部，用于防止来自患者的病毒向医务人员传播。 |
| 附件 | 产品技术要求 |
| 其他内容 | / |
| 备注 | 1. 本产品为新型冠状病毒肺炎应急防护用品，注册证有效期一年。 2. 注册人需在延续时按照补正通知要求完善生物相容性评价、产品有效期确认等相关资料。 3. 按照《医疗器械生产质量管理规范》的规定完善质量管理体系，加强产品原材料供应商审计和产品质量管理。 |

审批部门：安徽省药品监督管理局

批准日期：2020年07月02日

有效期至：2021年07月01日

医疗器械注册证（医用一次性防护服）

Registration Certificates (Medical Coveralls)

中华人民共和国医疗器械注册证

注册证编号：皖械注准 20202140060

| | |
|-------|--|
| 注册人名称 | 安徽天康医疗科技股份有限公司 |
| 注册人住所 | 安徽省天长市经济开发区纬一路 228 号 |
| 生产地址 | 安徽省天长市经济开发区纬一路 228 号 |
| 产品名称 | 医用一次性防护服 |
| 型号、规格 | 型号：连身式（含脚套和不含脚套）； 规格：160、165、170、175、180、185。 |
| 结构及组成 | 由连帽上衣和裤子组成。灭菌型产品经环氧乙烷灭菌，产品应无菌。 |
| 适用范围 | 供临床医务人员在工作时接触到的具有潜在感染性的患者血液、体液、分泌物等提供阻隔、防护用。 |
| 附件 | / |
| 其他内容 | / |
| 备注 | 1. 本产品为新型冠状病毒肺炎应急防护用品，注册证有效期一年。 2. 注册人需在延续注册时按照补正通知要求完善生物相容、产品有效期和包装研究资料。 3. 上市后加强产品原材料供应商管理和产品质量控制。 |

审批部门：安徽省药品监督管理局

批准日期：2020年03月05日

有效期至：2021年03月04日

医疗器械注册证（红外线额温计）

Registration Certificates (Infrared Forehead Thermometer)

中华人民共和国医疗器械注册证

注册证编号：皖械注准 20202070356

| | |
|-------|---|
| 注册人名称 | 安徽天康医疗科技股份有限公司 |
| 注册人住所 | 安徽省天长市经济开发区纬一路 228 号 |
| 生产地址 | 安徽省天长市经济开发区纬一路 228 号 |
| 代理人名称 | 不适用 |
| 代理人住所 | 不适用 |
| 产品名称 | 红外线额温计 |
| 型号、规格 | TKWQ-01 |
| 结构及组成 | 主要由 LCD 显示屏、红外探测头、PCBA 线路板、塑胶壳、按键、蜂鸣器（喇叭）组成。 |
| 适用范围 | 通过测量额头的热辐射来显示被测对象的体温。 |
| 附件 | 产品技术要求。 |
| 其他内容 | / |
| 备注 | 1. 本产品为新型冠状病毒肺炎应急防护用品，注册证有效期一年。 2. 注册人需在延续时按照补正通知要求完善生物相容性及产品临床评价资料等相关资料。 3. 按照《医疗器械生产质量管理规范》的规定完善质量管理体系，加强产品原材料供应商审计和产品质量管理。 |

审批部门：安徽省药品监督管理局

批准日期：2020 年 07 月 03 日
有效期至：2021 年 07 月 02 日



医疗器械生产备案凭证（医用隔离面罩、隔离衣、医用隔离眼罩）

Medical Device Filing Certificate (Medical Isolation Face-shield, Non-woven Isolation Gowns, Medical Isolation Goggles)

第一类医疗器械生产备案凭证

备案编号：皖滁食药监械生产备 20200002 号


| | | | | |
|--------|---|-----------------|------------|----|
| 企业名称 | 安徽天康医疗科技股份有限公司 | | | |
| 住 所 | 安徽省天长市经济开发区纬一路 228 号 | | | |
| 生产地址 | 安徽省天长市经济开发区纬一路 228 号 安徽省天长市经济开发区纬一路 228 号 | | | |
| 法定代表人 | 柏保东 | 企业负责人 | 柏保东 | |
| 生产范围 | 2002 分类目录 I 类: 6834-1- 医用纺织防护用品 2017 分类目录 I 类: 14-14- 医护人员防护用品 | | | |
| 生产产品列表 | 产品名称 | 产品备案号 | 登载日期 | 备注 |
| | 医用隔离面罩 | 皖滁械备 20200033 号 | 2020-02-11 | |
| | 隔离衣 | 皖滁械备 20200034 号 | 2020-02-11 | |
| | 医用隔离眼罩 | 皖滁械备 20200040 号 | 2020-03-09 | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |

备案部门（公章）：安徽省滁州市市场监督管理局
备案日期：2020 年 3 月 9 日



医疗器械生产备案凭证 (医用隔离鞋套)

Medical Device Filing Certificate(Medical Isolation Shoe Cover)



第一类医疗器械备案信息表

备案号：皖滁械备20200091号

| | |
|-----------|--|
| 备案人名称 | 安徽天康医疗科技股份有限公司 |
| 备案人社会信用代码 | 9134110071990801XY |
| 备案人注册地址 | 安徽省滁州市天长市经济开发区纬一路228号 |
| 生产地址 | 安徽省滁州市天长市经济开发区纬一路228号 |
| 代理人 | |
| 代理人注册地址 | |
| 产品名称 | 医用隔离鞋套 |
| 型号/规格 | S、M、L、XL |
| 产品描述 | 采用非织造复合布制成，有足够的强度和阻隔性能，非无菌提供。 |
| 预期用途 | 医务人员在医疗机构中使用，防止接触到具有潜在感染性的患者血液、体液、分泌物等，起阻隔、防护作用。 |
| 备注 | / |
| 备案单位和日期 | 滁州市市场监督管理局 备案日期：2020年08月27日 |
| 变更情况 | |

医疗器械产品出口销售证明

Certificate for Exportation of Medical Products

中华人民共和国
PEOPLE'S REPUBLIC OF CHINA
医疗器械产品出口销售证明
CERTIFICATE FOR EXPORTATION OF MEDICAL
PRODUCTS

证书编号：皖滁食药监械出 202000021 号
Certificate NO.：皖滁食药监械出 202000021 号

产品名称：口腔种植一次性使用输水管、一次性使用医用口罩、医用外科口罩、医用一次性防护服
Product(s)：Oral tubing、Disposable medical face mask、Surgical mask、Disposable medical coveralls

规格型号：见附件
Model：SEE ATTACHMENT

产品注册或备案凭证号：皖械注准 20192170111、皖械注准 20202140058、皖械注准 20202140059、皖械注准 20202140060
Registration certificate(s)：皖械注准 20192170111、皖械注准 20202140058、皖械注准 20202140059、皖械注准 20202140060

生产企业：安徽天康医疗科技股份有限公司
Manufacturer：Anhui Tiankang Medical Technology Co., Ltd

生产企业住所：安徽省天长市经济开发区纬一路 228 号
Address of manufacturer：No. 228 Weiyi Road, Economic Development Zone, Tianchang City, 239300 Anhui, China

生产许可或备案凭证号：皖食药监械生产许 20150052 号
Manufacturing License(s)：皖食药监械生产许 20150052 号

兹证明上述产品已准许在中国生产和销售。
This is to certify that the above products have been registered to be manufactured and sold in China.

证明有效期至：2021 年 03 月 04 日
This certification valid until: March 04th, 2021

备注：/
Remark：/



附件 (ATTACHMENT)

| 序号 | 产品名称 | Product(s) | 规格型号 | Model |
|----|--------------|------------------------------|---|--|
| 1 | 口腔种植一次性使用输水管 | Oral tubing | TKSSG-01 型、TKSSG-02 型、TKSSG-03 型 | TKSSG-01、TKSSG-02、TKSSG-03 |
| 2 | 一次性使用医用口罩 | Disposable medical face mask | 按口罩带的型式分为系带式和松紧带两种，按灭菌方式分为灭菌型和非灭菌型两种型号；按尺寸分为：17cm×8cm、17cm×9cm、17cm×9.5cm、17cm×10cm、17.5cm×8cm、17.5cm×9cm、17.5cm×9.5cm、17.5cm×10cm、18cm×8cm、18cm×9cm、18cm×9.5cm、18cm×10cm、19.5cm×8cm、19.5cm×9cm、19.5cm×9.5cm、19.5cm×10cm、12.5cm×9cm、14.5cm×9cm、16.5cm×9cm 十九个规格 | Different types: earloop mask、adjustable strap mask; Sterile and non-sterile Specifications: 17cm×8cm、17cm×9cm、17cm×9.5cm、17cm×10cm、17.5cm×8cm、17.5cm×9cm、17.5cm×9.5cm、17.5cm×10cm、18cm×8cm、18cm×9cm、18cm×9.5cm、18cm×10cm、19.5cm×8cm、19.5cm×9cm、19.5cm×9.5cm、19.5cm×10cm、12.5cm×9cm、14.5cm×9cm、16.5cm×9cm |
| 3 | 医用外科口罩 | Surgical mask | 按口罩带的型式分为系带式和耳挂式两种，按灭菌方式分为灭菌型和非灭菌型两种型号；按尺寸分为：12.5cm×9cm、14.5cm×9cm、16.5cm×9cm、17cm×8cm、17cm×9cm、17cm×9.5cm、17cm×10cm、17.5cm×8cm、17.5cm×9cm、17.5cm×9.5cm、17.5cm×10cm、18cm×8cm、18cm×9cm、18cm×9.5cm、18cm×10cm、19.5cm×8cm、19.5cm×9cm、19.5cm×9.5cm、19.5cm×10cm 十九个规格 | Different types: earloop mask、adjustable strap mask; Sterile and non-sterile Specifications: 12.5cm×9cm、14.5cm×9cm、16.5cm×9cm、17cm×8cm、17cm×9cm、17cm×9.5cm、17cm×10cm、17.5cm×8cm、17.5cm×9cm、17.5cm×9.5cm、17.5cm×10cm、18cm×8cm、18cm×9cm、18cm×9.5cm、18cm×10cm、19.5cm×8cm、19.5cm×9cm、19.5cm×9.5cm、19.5cm×10cm |
| 4 | 医用一次性防护服 | Disposable medical coveralls | 型号：连身式（含脚套和不含脚套）；规格：160、165、170、175、180、185。 | Type: One piece coveralls (with shoe cover; without shoe cover; Specification: 160、165、170、175、180、185 |

CE 注册证明 (I类)
CE Registration Certificate(Class I)

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMD/IV)
Formularnummer 00297923

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG
General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika
Form for Medical Devices except In Vitro Diagnostic Medical Devices

| | |
|---|---|
| Zuständige Behörde / Competent authority | |
| Code DE/CA61 | |
| Bezeichnung / Name Regierung von Oberbayern | |
| Staat / State Deutschland | Land / Federal state Bayern |
| Ort / City München | Postleitzahl / Postal code 80534 |
| Straße, Haus-Nr. / Street, house no. Maximilianstraße 39 | |
| Telefon / Phone +49-89-21760 | Telefax / Fax +49-89-21762914 |
| E-Mail / E-mail medizinprodukteanzeigenverfahren@reg-ob.bayern.de | |
| Anzeige / Notification | |
| Registrierdatum bei der zuständigen Behörde Registration date at competent authority 13.05.2020 | Registriernummer / Registration number DE/CA61/1M50/121 |
| Typ der Anzeige / Notification type S Erstanzeige / Initial notification E Änderungsanzeige / Notification of change E Widerrufsanzeige / Notification of withdrawal | |
| Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn | |
| Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG E Hersteller / Manufacturer S Bevollmächtigter / Authorised Representative E Einführer / Importer E Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG E Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV E Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG | |

- 1 -

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMD/IV)
Formularnummer 00297923

| | |
|--|---------------------------|
| Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market) | |
| Klasse / Class S I E I - steril / sterile E I - mit Messfunktion / with measuring function E I - steril und mit Messfunktion / sterile and with measuring function E IIa E IIb E III E III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 E Aktives implantierbares Medizinprodukt / Active implantable medical device E Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 | |
| App (Software auf mobilen Endgeräten) | E ja / yes S nein / no |
| Nummer(n) der Bescheinigung(en) / Certificate number(s) | |
| Handelsname des Produktes / Trade name of the device Single-use Medical Face Mask | |
| Produktbezeichnung / Name of device | |
| Nomenklaturcode / Nomenclature code 12-447 | |
| Nomenklaturbezeichnung / Nomenclature term Maske | |
| Kategoriecode / Category code 10 | |
| Kategorie / Category Produkte zum Einmalgebrauch | |
| Kurzbeschreibung deutsch / German short description | |
| Kurzbeschreibung englisch / English short description | |

- 4 -

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMD/IV)
Formularnummer 00297927

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG
General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika
Form for Medical Devices except In Vitro Diagnostic Medical Devices

| | |
|---|--|
| Zuständige Behörde / Competent authority | |
| Code DE/CA61 | |
| Bezeichnung / Name Regierung von Oberbayern | |
| Staat / State Deutschland | Land / Federal state Bayern |
| Ort / City München | Postleitzahl / Postal code 80534 |
| Straße, Haus-Nr. / Street, house no. Maximilianstraße 39 | |
| Telefon / Phone +49-89-21760 | Telefax / Fax +49-89-21762914 |
| E-Mail / E-mail medizinprodukteanzeigenverfahren@reg-ob.bayern.de | |

| | |
|---|---|
| Anzeige / Notification | |
| Registrierdatum bei der zuständigen Behörde Registration date at competent authority 24.06.2020 | Registriernummer / Registration number DE/CA61/1M50/203 |
| Typ der Anzeige / Notification type S Erstanzeige / Initial notification E Änderungsanzeige / Notification of change E Widerrufsanzeige / Notification of withdrawal | |
| Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn | |
| Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG E Hersteller / Manufacturer S Bevollmächtigter / Authorised Representative E Einführer / Importer E Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG E Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV E Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG | |

- 1 -

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMD/IV)
Formularnummer 00297927

Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)

| | |
|--|------------------------|
| Klasse / Class S I E I - steril / sterile E I - mit Messfunktion / with measuring function E I - steril und mit Messfunktion / sterile and with measuring function E IIa E IIb E III E III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 E Aktives implantierbares Medizinprodukt / Active implantable medical device E Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 | |
| App (Software auf mobilen Endgeräten) | E ja / yes S nein / no |
| Nummer(n) der Bescheinigung(en) / Certificate number(s) | |
| Handelsname des Produktes / Trade name of the device Non-woven Isolation Gowns | |
| Produktbezeichnung / Name of device | |
| Nomenklaturcode / Nomenclature code 15-037 | |
| Nomenklaturbezeichnung / Nomenclature term Kleidung, Isolation | |
| Kategoriecode / Category code 10 | |
| Kategorie / Category Produkte zum Einmalgebrauch | |
| Kurzbeschreibung deutsch / German short description | |
| Kurzbeschreibung englisch / English short description | |

- 4 -

ZOUSTECH

www.zoustech.eu

Zoustech, your trusted partner

Representing you in EU!

CONFIRMATION OF PRODUCT NOTIFICATION

This is to confirm that Zoustech S.L., has registered under the AEMPS (Spanish Agency for Medicines and Medical Devices), the following medical devices:

| Number in the contract | Product name in English | Product name in Spanish |
|------------------------|-------------------------|-------------------------|
| | Medical Coverall | Mono Médico |

Manufacturer: ANHUI TIANKANG MEDICAL TECHNOLOGY CO., LTD

Address: No. 228 Weiyi Road, Economic Development Zone, Tianchang City, 239300 Anhui, P.R. China,

Registered under number: RPS/786/2020 (See attached the electronic notification)

The Manufacturer has declared that these devices comply with the regulation including all the general safety and performance requirements.

Zoustech has complied with its commitment of registering the above mentioned devices under the AEMPS and will not have any other further obligation, compromise or responsibility.

13 May 2020

Mr. Rubén Valle Ibaseta
On behalf of
ZOUSTECH SL

ZOUSTECH S.L.
Pvo. Castellana, 141 - Planta 19
28049 Madrid - Spain
CIF: B87637591

INSCRITA EN EL REGISTRO MERCANTIL DE MADRID, TOMO 35086, FOLIO 147, HOJA M-630984, INSCRIPCIÓN 1

Envíos Telemáticos
Page 1 of 1

GOBIERNO DE ESPAÑA
MINISTERIO DE SANIDAD

Agencia Española de Medicamentos y Productos Sanitarios

Envíos Telemáticos

3.0.44

Registro de Responsables de Productos Sanitarios
Usuario: RUBÉN VALLE IBASETA
Desconectar

Registro de Responsables de Productos Sanitarios - RPS/786/2020

Datos de la notificación

Datos de registro

Nº Registro: Fecha Registro:

Datos del Responsable

Tipo de Responsable (*): Tipo de entidad:

CIF (*): Nombre (*):

Dirección (*):

Localidad (*):

Provincia (*): CP (*):

Teléfono (*): Fax:

e-mail (*): Web:

Datos del Fabricante

Nombre o Razón Social (*):

Dirección (*):

Localidad (*):

País (*): CP:

Teléfono (*): Fax:

e-mail (*): Web:

Datos de Productos Comunicados

Estatus (*):

Relación de Productos

Listado de Productos Sanitarios

2 filas, mostrando todas las filas

| Nombre Comercial | Tipo de Producto Estado del producto Acción |
|------------------|---|
| MONO MÉDICO | Clase I Primera Comunicación |

Comentarios

Agencia Española de Medicamentos y Productos Sanitarios
Parque Empresarial "Las Mercedes", Edif. 8, C/ Campezo 1 - 28022 MADRID | e-Mail: incidencias_aplicaciones@aemps.es

<https://enviotelematico-cc.aemps.es/enviotelematico/rps/solicitud.do?metodo=detalle...> 13/05/2020

www.tkmedical.com

23

24

安徽天康(集团)股份有限公司
安徽天康医疗科技股份有限公司

美国食品和药物管理局 (FDA) 注册列名
US FDA Listing Numbers

View Your Registered Facilities

Owner/Operator: [Blank]

Show 10 per page

| Registration Number | Registration Status | Registration Status Reason | Facility Name | Address | Expiration Date | Action |
|---------------------|---------------------|--|--|---|-----------------|---------------|
| 300750059 | Active | Registration changed from inactive to active | ANHUI TIANKANG MEDICAL PRODUCT CO., LTD. | NO 208, WEIYI ROAD ECONOMIC DEVELOPMENT ZONE, TIANCHANG, ANHUI 230000 CHINA | 2020-12-31 | [Action Icon] |

Showing 1 to 1 of 1 entries

View Selected Listing Details

Listing Number: 0374373
Listing Status: Active

Product Code: HOY
Product Name: Shield, eye, ophthalmic (including sunlamp protective eyewear and post-mydratic eyewear)

| Registration # | Registration Status | Registration Status Reason | Activities |
|----------------|---------------------|--|-----------------------|
| 300750059 | Active | Registration changed from inactive to active | Contract Manufacturer |

View Proprietary Names and Labeling

View Your Device Listings

Owner/Operator: 10042273

Show 10 per page

| Listing Number | Listing Status | Premarket Submission Number | Product Code(s) | Device Name | Action |
|----------------|----------------|-----------------------------|-----------------|--|---------------|
| 0374375 | Active | | 0EA | Non-surgical isolation gown | [Action Icon] |
| 0374373 | Active | | HOY | Shield, eye, ophthalmic (including sunlamp protective eyewear and post-mydratic eyewear) | [Action Icon] |
| 0373039 | Active | | KHA | MASK, SCAVENGING | [Action Icon] |
| 0367980 | Active | K191642 | FMP | Syringe, piston | [Action Icon] |
| 0367823 | Active | K191643 | FMI | Needle, hypodermic, single lumen | [Action Icon] |
| 0356013 | Active | K183540 | PNR | External Syringes with arterial specific connectors | [Action Icon] |
| 0293722 | Inactive | K162436 | NYL | Handpiece, air-powered, root canal irrigation | [Action Icon] |
| 0292623 | Inactive | | EDG | HEAT SOURCE FOR BLEACHING TEETH | [Action Icon] |
| 0292035 | Inactive | K151151 | FFA | Set, administration, intravascular | [Action Icon] |
| 0289372 | Active | K152808 | FMP MEG | Syringe, piston SYRINGE, ANTISTICK | [Action Icon] |

Showing 1 to 10 of 26 entries

View Selected Listing Details

Listing Number: 0373039
Listing Status: Active

Product Code: KHA
Product Name: MASK, SCAVENGING

| Registration # | Registration Status | Registration Status Reason | Activities |
|----------------|---------------------|--|-----------------------|
| 300750059 | Active | Registration changed from inactive to active | Contract Manufacturer |

View Proprietary Names and Labeling

View Selected Listing Details

Listing Number: 0374375
Listing Status: Active

Product Code: 0EA
Product Name: Non-surgical isolation gown

| Registration # | Registration Status | Registration Status Reason | Activities |
|----------------|---------------------|--|-----------------------|
| 300750059 | Active | Registration changed from inactive to active | Contract Manufacturer |

View Proprietary Names and Labeling

View Proprietary Names and Labeling

Listing Number: 0373039

| Proprietary Name | Confidential | Device labeled for use | Device Identifier | Label |
|-----------------------------|--------------|------------------------|-------------------|-------|
| Single use medical facemask | Y | | | |

Close and Return

俄罗斯海关联盟 EAC 认证
EAC certification of Russian customs union

官方发布防疫物资出口企业白名单
Official Release White List of Anti-epidemic Supplies Export Companies



ISO 9001:2015 质量管理体系认证证书
ISO 9001:2015 Quality Management System Certificate



质量管理体系认证证书

注册号: U006619Q0079R0M

兹证明

安徽天康医疗科技股份有限公司

统一社会信用代码: 9134110071990801XY
注册/办公/生产地址: 中国·安徽省·天长市经济开发区纬一路 228 号

质量管理体系符合标准
ISO 9001:2015

认证范围如下:

一次性使用无菌自毁型固定剂量疫苗注射器(带针), 一次性使用无菌注射器(带针), 一次性使用医用口罩、医用外科口罩、医用一次性防护服、隔离衣、医用隔离面罩、医用防护口罩的生产, 一次性手术衣、喂液管的销售

换证日期: 2020 年 10 月 10 日
本证书有效期自 2019 年 3 月 18 日至 2022 年 3 月 17 日
标准适用性细节参见组织的质量手册
认证范围涉及法律法规要求的行政许可、资质许可、强制性认证的, 证书与资质共同使用有效。
在正常接受年度审核的情况下, 与年度监督保持通知一并使用有效。
本证书信息可在国家认证认可监督管理委员会官方网站 (www.cnca.gov.cn) 上查询。



华夏认证中心有限公司
中国北京市海淀区北四环中路 211 号太极大厦
<http://www.ccci.com.cn>

总经理: 

颁证日期: 2019 年 3 月 18 日




0066



QUALITY MANAGEMENT SYSTEM CERTIFICATE

Registration No. U006619Q0079R0M

This is to certify that the quality management system of

Anhui Tiankang Medical Technology Co., Ltd.

social credit code : 9134110071990801XY
Registration/Office/Production Address: No. 228 Weyi Road, Economic Development Zone, Tianchang City, Anhui, P.R. China

is in conformity with
ISO 9001:2015

This certificate is valid for the following scope:

The production of Auto-disable Syringes for Fixed-dose Immunization (with needle), Disposable Syringes (with needle), Single-use Medical Face Mask, Surgical Mask, Medical Coveralls, Non-woven Isolation Gowns, Medical Isolation Face-shield, Protective Face, Mask for Medical Use, The sale of Single-use Surgical Gown, Feeding Tube

The certificate changed on October 10, 2020
This certificate is valid from March 18, 2019 to March 17, 2022
The applicability of the standard are detailed in the organization's Quality Manual.
The scope which needs administrative permission shall be in accordance with valid license.
In the case that the organization regularly receives surveillance assessments, the certificate shall be valid when used together with the Notice for Maintenance of Use of Certificates and Logos.
Information about the certificate can be queried on the official website of CNCA (www.cnca.gov.cn).



China Certification Center, Inc.
No. 211 Commercial Building, No. 211 Beisihuan Zhonglu,
Haidian District, Beijing, P.R. China
<http://www.ccci.com.cn>

General Manager: 

Date of Issue: March 18, 2019




0066

ISO 13485:2016 & EN ISO 13485:2016 质量管理体系认证证书
The Certification of ISO 13485:2016 & EN ISO 13485:2016 Quality Management System Certificate

CE 认证证书 (I 类灭菌)
CE Certificate(Class Is)




By Royal Charter

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016 & EN ISO 13485:2016

This is to certify that: Anhui Tiangkang Medical Technology Co., Ltd. 安徽天康医疗科技股份有限公司
No. 228 Weiyi Road 中国
Economic Development Zone 安徽省
Tianchang City 天长市
Anhui 经济开发区
239300 纬一路228号
China 邮编: 239300

Holds Certificate No: **MD 730170**

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 & EN ISO 13485:2016 for the following scope:

Production and Distribution of Single-use Medical Face Mask, Medical Coveralls, Non-woven Isolation Gowns, Medical Isolation Goggles.
一次性使用医用口罩, 医用一次性防护服, 隔离衣, 医用隔离眼罩的生产和销售。

Gary E Slack

For and on behalf of BSI: **Gary E Slack, Senior Vice President - Medical Devices**

Original Registration Date: 2020-07-29 Effective Date: 2020-07-29
Latest Revision Date: 2020-07-29 Expiry Date: 2023-07-28

Page: 1 of 1

...making excellence a habit™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract. An electronic certificate can be authenticated [online](https://www.bsi-global.com/ClientDirectory). Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +86 10 8507 3000.

Information and Contact: BSI, John M. Keynesplein 9, 1066 EP Amsterdam The Netherlands. Tel: +31 (0) 20 3460 780
BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A Member of the BSI Group of Companies.




By Royal Charter

EU Quality Assurance Certificate

Regulation (EU) 2017/745, Annex XI Part A

MDR 730169 R000

Manufacturer: Anhui Tiangkang Medical Technology Co., Ltd.

Address:
No. 228 Weiyi Road
Economic Development Zone
Tianchang City
Anhui
239300
China

Single Registration Number: Not Available

EU Authorised Representative: MedPath GmbH

Address:
Mies-van-der-Rohe-strasse8,
80807 Munich,
Germany

Scope: See attached **Device Schedule**

On the basis of our examination of the quality system in accordance with Regulation (EU) 2017/745, Annex XI part A, the quality system meets the requirements of the Regulation. For the placing on the market of Class III and Class IIb devices an additional Annex X certificate is required.

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Gary E Slack

Gary E Slack, Senior Vice President Medical Devices



First Issued: **2020-08-20** Date: **2020-08-20** Expiry Date: **2025-08-19**

...making excellence a habit™

Page 1 of 3

Validity of this certificate is conditional on the Manufacturer's quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body. This certificate was issued electronically and is bound by the conditions of the contract.

NB Contact: BSI Group The Netherlands B.V., Say Building, John M. Keynesplein 9, 1066 EP, Amsterdam, Netherlands. Tel: +31 (0) 20 346 07 80
Corporate Contact: BSI Group Assurance Limited, registered in England under number 05435540 at 389 Chiswick High Road, London, W4 4AL, UK.
A Member of the BSI Group of Companies.

EU Quality Assurance Certificate
Regulation (EU) 2017/745, Annex XI Part A

MDR 730169 R000

Device Schedule: Class III and Class IIb devices

Device Schedule: Class IIa, Custom-made and other devices

| Device(s) | Risk Classification |
|------------------------------|---------------------|
| Single-use Medical Face Mask | Class Is |
| Medical Coveralls | Class Is |

For Class Is devices, the Notified Body conformity assessment is limited to the aspects relating to establishing, securing and maintaining sterile conditions.

First Issued: **2020-08-20**

Date: **2020-08-20**

Expiry Date: **2025-08-19**

...making excellence a habit.[™]

Page 2 of 3

Validity of this certificate is conditional on the Manufacturer's quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

NB Contact: BSI Group The Netherlands B.V., Say Building, John M. Keynesplein 9, 1066 EP, Amsterdam, Netherlands. Tel: + 31 (0) 20 346 07 80
Corporate Contact: BSI Group Assurance Limited, registered in England under number 05435540 at 389 Chiswick High Road, London, W4 4AL, UK.
A Member of the BSI Group of Companies.




EU Quality Assurance Certificate
Regulation (EU) 2017/745, Annex XI Part A

MDR 730169 R000

Certificate History
(References to applicable Common Specifications, Harmonized Standards complied with, and the relevant test and audit reports that support any of the below certificate changes may be requested from Certificate.Verification@bsigroup.com)

| Date | Reference number | Action |
|---------|------------------|--------------|
| Current | 3220023 | First Issue. |

First Issued: **2020-08-20**

Date: **2020-08-20**

Expiry Date: **2025-08-19**


...making excellence a habit.[™]

Page 3 of 3

Validity of this certificate is conditional on the Manufacturer's quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

NB Contact: BSI Group The Netherlands B.V., Say Building, John M. Keynesplein 9, 1066 EP, Amsterdam, Netherlands. Tel: + 31 (0) 20 346 07 80
Corporate Contact: BSI Group Assurance Limited, registered in England under number 05435540 at 389 Chiswick High Road, London, W4 4AL, UK.
A Member of the BSI Group of Companies.

CE 认证证书 (PPE 认证)
CE Certificate(PPE Certificate)




CERTYFIKAT BADANIA TYPU UE (MODUŁ B)
EU TYPE-EXAMINATION CERTIFICATE (MODULE B)

Nr
No. CW/PPER/9/09/2020

ZAŚWIADCZA SIĘ,
że Polski Rejestr Statków S.A. (PRS) przeprowadził procedurę badania typu wymienionego niżej wyrobu i stwierdził jego zgodność z wymaganiami określonymi w załączniku V do Rozporządzenia Parlamentu Europejskiego i Rady (UE) 2016/425 (PPE) w sprawie środków ochrony indywidualnej oraz uchylenia dyrektywy Rady 89/686/EEG, ze zmianami.

THIS IS TO CERTIFY
that Polski Rejestr Statków S.A. (PRS) did undertake the EU type-examination procedure for the product identified below which was found to be in compliance with the requirements of Annex V to the Regulation (EU) 2016/425 of the European Parliament and of the Council of 9 March 2016 on personal protective equipment and repealing Council Directive 89/686/EEC, as amended.

| | |
|--|--|
| Wnioskodawca Applicant | Anhui Tiankang Medical Technology Co., Ltd. No.228, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province 239300, China |
| Producent Manufacturer | Anhui Tiankang Medical Technology Co., Ltd. No.228, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province 239300, China |
| Typ wyrobu Product type | Sprzęt ochrony dróg oddechowych. Półmaski filtrujące do ochrony przed cząstkami. Respiratory protective devices. Filtering half masks to protect against particles. |
| Opis wyrobu Product description | Półmaska filtrująca FFP2 NR, Model: TK-3 Particle Filtering Half Mask FFP2 NR, Model: TK-3 |
| Zastosowane normy Specified standards | PN-EN 149+A1:2010 EN 149:2001+A1:2009 |


Niniejszy certyfikat pozostaje ważny do czasu unieważnienia przy zachowaniu warunków uznania (patrz str. 2).
This certificate remains valid unless cancelled or revoked, provided the approval conditions (see page 2) are complied with.

Data ważności
Expiry date: 2025-09-02



PRS
1934

Zastępca Dyrektora Pionu Certyfikacji
Certification Division Director



Przemysław Gałka



Nr jednostki notyfikowanej
No. of notified body
1463

NOTIFIED BODY
NO.1463

Polski Rejestr Statków S.A.
ul. Gen. Józefa Hallera 126
80-416 Gdańsk, Poland

tel. (+48) (58) 346 17 00
fax (+48) (58) 341 77 69
e-mail: dc@prs.pl
www: http://www.prs.pl/

Gdańsk, 2020-09-03

Form. 8/PCW-01/PPER
2020-03-26

1/2

CW/PPER/9/09/2020

Wykaz dokumentacji
List of documents

- Instrukcja użytkowania - zatwierdzona przez PRS dnia 2020-09-02.
- Ocena ryzyka - zatwierdzona przez PRS dnia 2020-09-02.
- Rysunek półmaski - model: TK-3 - zatwierdzony przez PRS dnia 2020-09-02.
- Raport z badań nr [2020] WSZ FHL NO. 6188 wydany przez Jiangsu Guojian Testing Technology Co., Ltd w dniu 2020-08-04.
- Sprawozdanie z przeglądu PRS nr CW/KKr/PPER/19/2020 z dnia 2020-09-03.

1. Instruction for use - approved by PRS on 2020-09-02.
2. Risk analysis - approved by PRS on 2020-09-02.
3. Assembly drawing for model: TK-3 - approved by PRS on 2020-09-02.
4. Test report no. [2020] WSZ FHL NO. 6188 issued by Jiangsu Guojian Testing Technology Co., Ltd on 2020-08-04.
5. PRS Survey Report No. CW/KKr/PPER/19/2020 dated on 2020-09-03.

Miejsca produkcji
(inne niż podane na stronie 1)
Places of production
(different than given on page 1)

Ograniczenia uznania
Approval limitations

- Dane techniczne:
- półmaska z regulowanym metalowym klipsem na nos,
- półmaska wykonana z 5-warstwowej włókniny z filtrem z tkaniny,
- wymiary: 160 x 110 mm,
- kolor: biały.
- Półmaski filtrujące przeznaczone do jednorazowego użytku.
- Dokumentacja techniczna zatwierdzona w języku angielskim.

- Specifications:
- half mask with adjustable nose clip,
- half mask made with 5-layers non-woven fabric with melt-blown fabric filter,
- size: 160 x 110 mm,
- color: white.
- Filtering half mask shall not be used for more than one shift.
- Technical documentation approved in English.

Warunki uznania
Approval conditions

- Niniejszy certyfikat straci ważność po wprowadzeniu zmian lub modyfikacji w wyrobie bez uprzedniego uzgodnienia z PRS.
This certificate becomes invalid after changes or modifications to the product without prior agreement with PRS.
- Znak zgodności może być umieszczony na uznanym wyrobie oraz może być wystawiona deklaracja zgodności tylko pod warunkiem, że łącznie z badaniem typu UE zostanie przeprowadzona ocena zgodności produkcji pod nadzorem jednostki notyfikowanej, według załącznika VII lub VIII wymienionego wyżej rozporządzenia.
The Mark of Conformity may only be affixed to the above type approved product and a manufacturer's Declaration of Conformity issued provided the production is assessed under surveillance of a notified body according to Annex VII or VIII of the a/m Regulation.

Form. 8/PCW-01/PPER
2020-03-26

2/2



AC 114

CERTYFIKAT ZGODNOŚCI Z TYPEM W OPARCIU O WEWNĘTRZNĄ KONTROLĘ PRODUKCJI ORAZ NADZOROWANE KONTROLE PRODUKTU W LOSOWYCH ODSTĘPACH CZASU (Moduł C2)

CONFORMITY TO TYPE CERTIFICATE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECKS AT RANDOM INTERVALS (Module C2)

Nr CW/PPER/1/10/2020 Okres objęty certyfikatem 2020-10-01 – 2021-09-30
 No. CW/PPER/1/10/2020 Period covered by the certificate 2020-10-01 – 2021-09-30

Dokumenty odniesienia: Rozporządzenie UE 2016/425 dotyczące środków ochrony indywidualnej (PPE), załącznik VII
 General reference documents: Regulation (EU) 2016/425 on personal protective equipment (PPE), Annex VII

Posiadacz certyfikatu: **Anhui Tiangkang Medical Technology Co., Ltd.**
 Certificate holder: No.228, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province 239300, China.

| Wyrób Product | Certyfikat badania typu UE EU Type-examination certificate | Normy zharmonizowane/Specyfikacje Harmonised standards/Specifications |
|---|---|--|
| Półmaska filtrująca FFP2 NR, model: TK-3 Particle Filtering Half Mask FFP2 NR, Model: TK-3 | CW/PPER/9/09/2020 | PN-EN 149+A1:2010 EN 149:2001+A1:2009 |

A Roczna ocena zgodności wyrobów z normą/specyfikacją i badanym typem
Annual assessment of products compliance with standard/specification and type-examined

1 Miejsca i daty wizyt
 Visit locations and dates: Anhui Tiangkang Medical Technology Co., Ltd

2a Wyboru dokonał (imię, nazwisko)
 Selection carried out by (Name): Krzysztof Kirysiuk
 Związek z jednostką notyfikowaną
 Relationship to notified body: Ekspert Biura Certyfikacji Wyrobów i Osób
 Products and Persons Certification Bureau Expert

2b Przedstawiciel firmy (imię, nazwisko)
 Company representative (Name): Baodong Bai
 Stanowisko
 Position: Legal Representative

3 Związek pomiędzy wizytowaną firmą a posiadaczem certyfikatu badania typu UE
 Relationship of company visited to EU type-examination certificate holder

Posiadacz certyfikatu
 Certificate holder Miejsce produkcji
 Production site Inne miejsce produkcji
 Secondary production site Importer
 Importer Dystrybutor
 Distributor

Sprzedaż detaliczna
 Retail outlet Europejskie biuro firmy
 European office of the company Inny:
 Other:

Wykaz środków ochrony indywidualnej
 List of personal protection equipment Dostępny
 Available Niedostępny
 Not available

Wybór próbki
 Sample selection Wybrano – Nr egz./partii:
 Selected – lot/batch No. 001/NA/CL/714/2020-009/NA/CL/714/2020;
 001/BR/714/2020-009/BR/714/2020 Nie wybrano
 Not selected

4 Wybór próbki
 Sample selection Prawidłowy
 Correct Nieprawidłowy
 Incorrect Wyniki badań
 Result of tests Pozytywne
 Positive Negatywne
 Negative

5 Wybór próbki i badania wykazały zgodność z przywołanymi normami/specyfikacjami i badanym typem
 Sample selection and testing demonstrated compliance with the reference standards/specifications and type-examined Tak
 Yes Nie
 No

CE Nr jednostki notyfikowanej
 No. of notified body: 1463
 Polski Rejestr Statków S.A.
 al. Gen. Józefa Hallera 126
 80-416 Gdańsk, Poland
 tel. (+48) (58) 346 17 00
 fax (+48) (58) 346 03 92
 e-mail: mailbox@prs.pl
 www: http://www.prs.pl/

Form. 10 PCW-01/PPER
 2020-03-26

1/2

CW/PPER/1/10/2020

B Roczna ocena niejednorodności produkcji
Annual assessment of production non-homogeneity

- 1 Zastosowana metoda przy dokonaniu oceny
 Method employed to perform assessment
- Inspekcja procesu produkcyjnego i zapisów z prób
 On-site review of production and test records
- Audit kontroli procesu produkcyjnego
 On-site audit of production control
- Ocena niejednorodności produkcji poprzez ocenę jednej dużej próbki
 Production non-homogeneity assessed by selection of a single, large sample
- Ocena niejednorodności produkcji poprzez ocenę próbek w ciągu roku
 Production non-homogeneity assessed by assessment of samples throughout the year

2a Ocenę przeprowadził (imię, nazwisko)
 Assessment carried out by (Name): ----

Związek z jednostką notyfikowaną
 Relationship to notified body: ----

2b Przedstawiciel firmy (imię, nazwisko)
 Company representative (Name): ----

Stanowisko
 Position: ----

3 Na podstawie przeprowadzonej oceny stwierdzono, że proces produkcyjny jest jednorodny
 On the basis of the assessment, it has been concluded the production is homogeneous Tak
 Yes Nie
 No

C Podsumowanie
Conclusion

Uzasadnienie niezgodności
 Justification of non-conformities

Nie było żadnych niezgodności / There were no non-conformities.

Wnioski jednostki notyfikowanej
 Conclusions of notified body

Środek ochrony osobistej jest kompatybilny z typem określonym w certyfikacie badania typu UE.
 Personal protective equipment is compatible with the type defined in the EC type-examination certificate.

Uwagi
 Remarks

1. Półmaska filtrująca przeznaczona do jednorazowego użytku.
2. Dokumentacja techniczna zatwierdzona w języku angielskim.
3. Produkt ten nie może być stosowany jako maska przeciwgazowa w środowisku toksycznym.
4. Półmaska filtrująca nie jest przeznaczona do użytkowania medycznego i chirurgicznego.

1. Filtering Half Mask shall not be used for more than one shift.
2. Technical documentation in English approved.
3. This product can not be used as a gas mask in a toxic environment.
4. Filtering Half Mask can not be used for medical or surgical purposes.

D Załączniki
Attachments

Sprawozdania z wizyty Nr
 Visit reports No. CW/KKr/PPER/26/2020

Sprawozdania z badań Nr
 Test reports No. Raport z badań nr CL/WBO/64/2020 wydany przez Laboratorium Badawcze PRS S.A. z dnia 2020-09-23.
 Test report no. CL/WBO/64/2020 issued by PRS S.A. Testing Laboratory dated on 2020-09-23.

Ogólna ocena z rocznego nadzoru
Overall assessment of the annual surveillance

Pozytywna
 Positive Negatywna
 Negative



Zastępca Dyrektora Biura Certyfikacji
 Certification Division Deputy Director

Przemysław Gałka

Gdańsk, 2020-10-01

Form. 10 PCW-01/PPER
 2020-03-26

2/2

澳大利亚治疗用品注册证书

Australian Register of Therapeutic Goods Certificate



医用一次性防护服（非灭菌）
Medical Coveralls(Non-sterile)

检验报告

报告编号: AH2020-QZC-00113

检品名称: 医用一次性防护服

检验目的: 注册检验

安徽省食品药品检验研究院



ZLJL-165-04

安徽省食品药品检验研究院 检验报告首页

报告编号: AH2020-QZC-00113

共 4 页 第 1 页

| | | | |
|-------|---|---------|-----------------------|
| 样品名称 | 医用一次性防护服 | 样品编号 | AH2020-QZC-00113 |
| | 送样(√) 抽样() | | |
| 商 标 | Tkmd | 型号规格 | 175 非灭菌 |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 注册检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号/批号 | 批 200101 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200120 |
| 抽样单位 | / | 样品数量 | 24个 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/02/18 | 检验日期 | 2020/02/18~2020/02/26 |
| 检验项目 | 2.1-2.8、2.9.1 | | |
| 检验依据 | 安徽天康医疗科技股份有限公司《医用一次性防护服》产品技术要求 | | |
| 检验结论 | 被检样品符合安徽天康医疗科技股份有限公司《医用一次性防护服》产品技术要求的要求。 (检验报告专用章或检验单位公章) 签发日期: 2020/2/26 | | |
| 备注 | 报告中“—”表示不适用项,“/”表示空白项。 | | |

批 准: 审 核: 主 检:

职 务: 授权签字人





检验报告

报告编号: AH2020-QZX-00053

检品名称: 一次性使用医用口罩

检验目的: 委托检验(生物学评价)

安徽省食品药品检验研究院



ZLJL-165-13

安徽省食品药品检验研究院 生物相容性试验报告

报告编号: AH2020-QZX-00053

共 3 页 第 1 页

| | | | | |
|-------|---|--------|---------|----------------------|
| 检品名称 | 一次性使用医用口罩 | | 样品编号 | AH2020-QZX-00053 |
| | 送样 (√) | 抽样 () | | |
| 商 标 | Tkmd | | 型号规格 | 17.5cm×9.5cm (灭菌型) |
| 委托方 | 安徽天康医疗科技股份有限公司 | | 检验类别 | 委托检验(生物学评价) |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | | 产品编号/批号 | 批200101 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | | 抽样单编号 | / |
| 受检单位 | / | | 生产日期 | 20200115 |
| 抽样单位 | / | | 样品数量 | 36个 |
| 抽样地点 | / | | 抽样基数 | / |
| 抽样日期 | / | | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/03/02 | | 检验日期 | 2020/03/02~2020/6/29 |
| 检验项目 | 细胞毒性、致敏、皮肤反应 | | | |
| 检验依据 | GB/T 16886.5-2017、GB/T 16886.10-2017 | | | |
| 检验结论 | 所检项目参照GB/T 16886.5-2017、GB/T 16886.10-2017标准检验，结果合格。 (检验报告专用章或检验单位公章) 签发日期: 2020/6/29 | | | |
| 备 注 | 1) 报告中的“—”表示此项不适用，报告中“/”表示此项空白。 | | | |

批 准: 章业兵 审 核: 孙必文 主 检: 周 莹
 职 务: 授权签字人



检验报告

报告编号: AH2020-QSJ-00482

检品名称: 医用外科口罩(灭菌)

检验目的: 委托检验

安徽省食品药品检验研究院



ZLJL-165-04

安徽省食品药品检验研究院 检验报告首页

报告编号: AH2020-QSJ-00482

共 3 页 第 1 页

| | | | |
|-------|---|---------|-----------------------|
| 样品名称 | 医用外科口罩(灭菌) | 样品编号 | AH2020-QSJ-00482 |
| | 送样(√) 抽样() | | |
| 商 标 | Tkmd | 型号规格 | 灭菌型/耳挂式 17.5cm×9.5cm |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 委托检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号/批号 | 批200501 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200520 |
| 抽样单位 | / | 样品数量 | 120个 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/06/04 | 检验日期 | 2020/06/04~2020/06/30 |
| 检验项目 | 4.1-4.8、4.9.2、4.10(除4.2) | | |
| 检验依据 | YY 0469-2011《医用外科口罩》 | | |
| 检验结论 | 所检项目符合YY 0469-2011《医用外科口罩》的要求。 (检验报告专用章或检验单位公章) 签发日期: 2020/6/30 | | |
| 备注 | 报告中“—”表示不适用项,“/”表示空白项。 | | |

批 准: 审 核: 主 检:
 职 务: 授权签字人



检验报告

报告编号: AH2020-QZX-00051

检品名称: 医用外科口罩

检验目的: 委托检验(生物学评价)

安徽省食品药品检验研究院



ZLJL-165-13

安徽省食品药品检验研究院 生物相容性试验报告

报告编号: AH2020-QZX-00051

共 3 页 第 1 页

| | | | | |
|-------|---|--------|---------|------------------------|
| 检品名称 | 医用外科口罩 | | 样品编号 | AH2020-QZX-00051 |
| | 送样 (√) | 抽样 () | | |
| 商 标 | Tkmd | | 型号规格 | 18cm×9.5cm (±5%) (灭菌型) |
| 委托方 | 安徽天康医疗科技股份有限公司 | | 检验类别 | 委托检验(生物学评价) |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | | 产品编号/批号 | 批200101 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | | 抽样单编号 | / |
| 受检单位 | / | | 生产日期 | 20200116 |
| 抽样单位 | / | | 样品数量 | 36个 |
| 抽样地点 | / | | 抽样基数 | / |
| 抽样日期 | / | | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/03/02 | | 检验日期 | 2020/03/02~2020/6/30 |
| 检验项目 | 细胞毒性、致敏、皮内反应 | | | |
| 检验依据 | GB/T 16886.5-2017、GB/T 16886.10-2017 | | | |
| 检验结论 | 所检项目参照GB/T 16886.5-2017、GB/T 16886.10-2017标准检验,结果见报告第2页。 (检验报告专用章或检验单位公章) 签发日期: 2020/6/30 | | | |
| 备 注 | 1) 报告中的“—”表示此项不适用,报告中“/”表示此项空白。 | | | |

批 准: 章业兵 审 核: 孙治云 主 检: 周 莹
 职 务: 授权签字人

医用一次性防护服（灭菌级）
Medical Coveralls(Sterile)

检验报告

报告编号：AH2020-QZC-00112

检品名称：医用一次性防护服

检验目的：注册检验

安徽省食品药品检验研究院



ZLJL-165-04

安徽省食品药品检验研究院 检验报告首页

报告编号：AH2020-QZC-00112

共 4 页 第 1 页

| | | | |
|-------|---|---------|-----------------------|
| 样品名称 | 医用一次性防护服 | 样品编号 | AH2020-QZC-00112 |
| | 送样(√) 抽样() | | |
| 商 标 | Tkmd | 型号规格 | 175 灭菌级 |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 注册检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号/批号 | 批 200101 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200120 |
| 抽样单位 | / | 样品数量 | 24个 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/02/18 | 检验日期 | 2020/02/18~2020/03/04 |
| 检验项目 | 2.1-2.8、2.9.2、2.10 | | |
| 检验依据 | 安徽天康医疗科技股份有限公司《医用一次性防护服》产品技术要求 | | |
| 检验结论 | 被检样品符合安徽天康医疗科技股份有限公司《医用一次性防护服》产品技术要求的要求。 (检验报告专用章或检验单位公章) 签发日期：2020/3/4 | | |
| 备注 | 报告中“—”表示不适用项，“/”表示空白项。 | | |



批 准： 审 核： 主 检：

职 务： 授权签字人



检验报告

报告编号: AH2020-QZX-00052

检品名称: 医用一次性防护服

检验目的: 委托检验(生物学评价)

安徽省食品药品检验研究院



ZLJL-165-13

安徽省食品药品检验研究院 生物相容性试验报告

报告编号: AH2020-QZX-00052

共 3 页 第 1 页

| | | | | |
|-------|---|--------|---------|----------------------|
| 检品名称 | 医用一次性防护服 | | 样品编号 | AH2020-QZX-00052 |
| | 送样 (√) | 抽样 () | | |
| 商 标 | Tkmd | | 型号规格 | 175 (灭菌型) |
| 委托方 | 安徽天康医疗科技股份有限公司 | | 检验类别 | 委托检验(生物学评价) |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | | 产品编号/批号 | 批200101 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | | 抽样单编号 | / |
| 受检单位 | / | | 生产日期 | 20200120 |
| 抽样单位 | / | | 样品数量 | 8个 |
| 抽样地点 | / | | 抽样基数 | / |
| 抽样日期 | / | | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/03/02 | | 检验日期 | 2020/03/02~2020/6/29 |
| 检验项目 | 细胞毒性、致敏、皮内反应 | | | |
| 检验依据 | GB/T 16886.5-2017、GB/T 16886.10-2017 | | | |
| 检验结论 | 所检项目参照GB/T 16886.5-2017、GB/T 16886.10-2017标准检验,结果见报告第2页。 (检验报告专用章或检验单位公章) 签发日期: 2020/6/29 | | | |
| 备 注 | 1) 报告中的“—”表示此项不适用, 报告中“/”表示此项空白。 | | | |

批 准: 章业兵 审 核: 孙冰云 主 检: 周 莹
职 务: 授权签字人



Test Report (Electronic version)

Verification Website: www.gtcc.net.cn
Verification Code: BXPk-0385-54

No:20R003207 Issue Date: 2020-07-01

Applicant: ANHUI TIANKANG MEDICAL TECHNOLOGY CO., LTD.
Address: NO. 228 WEIYI ROAD ECONOMIC DEVELOPMENT ZONE TIANCHANG CITY.

Information confirmed by applicant:

Medical coveralls
Quantity: 8 pieces
Brand: TKMD
Lot number: 200502
Production date: 20200502
Model: 185
Manufacture's name: ANHUI TIANKANG MEDICAL TECHNOLOGY CO., LTD.

Standard Adopted:

Date Received/Date Test Started: 2020-06-20

Conclusion:

Resistance to penetration by synthetic blood ---

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement "-"-No comment

Remark:
This report is the english translation version of the report 20R003206.
All the tested items are tested under the standard condition (except for indication).
Copies of the report are valid only re-stamped.
The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:
ZiShan Guo Senior Engineer

ZiShan Guo



Test Report (Electronic version)

No: 20R003207





Test Report (Electronic version)

Verification Website: www.gttc.net.cn
Verification Code: LTAW-6864-44

No: 20R003209 Issue Date: 2020-07-01

Applicant: ANHUI TIANKANG MEDICAL TECHNOLOGY CO., LTD.
Address: NO. 228 WEIYI ROAD ECONOMIC DEVELOPMENT ZONE TIANCHANG CITY.

Information confirmed by applicant:

Medical coveralls
Quantity: 8 pieces
Brand: TKMD
Lot number: 200502
Production date: 20200502
Model: 185
Manufacturer's name: ANHUI TIANKANG MEDICAL TECHNOLOGY CO., LTD.

Standard Adopted:

EN 14126:2003/AC:2004 <Protective clothing - Performance requirements and tests methods for protective clothing against infective agents>, Client Requirement

Date Received/Date Test Started: 2020-06-20

Conclusion:

| | |
|--|-----|
| The resistance to dry microbial penetration | M |
| The resistance to wet bacterial penetration[YY/T 0506.6-2009] | --- |
| The resistance to wet bacterial penetration[EN ISO 22610:2006] | --- |
| Resistance to penetration by blood-borne pathogens | M |

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement "----"-No comment

Remark:

This report is the english translation version of the report 20R003208.
The decision indicators are derived from the standard required by client (EN 14126:2003/AC:2004).Our inspection capacity authorized by CMA covers the inspection items EN 14126:2003/AC:2004.
All the tested items are tested under the standard condition (except for indication).
Copies of the report are valid only re-stamped.
The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:
ZiShan Guo Senior Engineer

ZiShan Guo



Test Report (Electronic version)

No: 20R003209



儿童卫生口罩
Children Hygiene Mask

检验报告
TEST REPORT

报告编号(No.): GS20000434-1 第 1 页共 6 页 Page 1of6

| | | | |
|---|---|-----------------------|---------------------|
| 委托方 Applicant | 安徽天康医疗科技股份有限公司 Anhui Tiankang Medical Technology Co.,Ltd | | |
| 生产方 Manufacture | 安徽天康医疗科技股份有限公司 Anhui Tiankang Medical Technology Co.,Ltd | | |
| 客户认定 样品信息 Submitted Sample Description | 商标 Trade Mark: — 样品名称 Sample Name: 儿童卫生口罩 Children hygiene Mask. 色号 Color: — 数量 Quantity: 1, 描述 Description: 白色口罩 white mask 产品等级 Sample Grade: —, 安全类别 Safety Category: — 声明款号 Style No.: — | | |
| 检验类型 Test Type | 委托检验 Requested Examination | 来样日期 Date Received | 签发时间 Date Issued |
| | | 2020-07-14 | 2020-07-30 |
| 判定依据 Performance Standard | GB/T 38880-2020 儿童口罩技术规范 GB/T 38880-2020 Technical specification of children mask | | |

批准 Approved by

Wang Baojun (Director)

委托检验仅对来样负责, 不承担其他连带责任。This report is valid only for the samples delivered by clients. 对检验报告若有异议, 应于收到报告之日起 15 日内向检验单位提出, 逾期不再受理。If there is any objection concerning the report, it is required that the objection should be put forth to the center within 15 days from the reception date of the report.

备注 Notes: 除非客户要求, 本报告检测结果及符合性判定不考虑测量结果的不确定度。Measurement uncertainty is not considered on the results and conformity assessment in the report unless required by the customer.
*暂未申请认可。*The test isn't requested for confirmation temporarily.
样品和数据采用 CF20000321-1

中纺标检认证股份有限公司
中国商业联合会轻纺海品质量监督检测中心
进出口商品检验鉴定机构
国家纺织制品质量监督检验中心

Chinatesta Textile Testing & Certification Services Tel: 010-65002813 400-066-0486
Add: No.3 Yanjingli Middle Street, Chaoyang District, Beijing 10025, China Fax: 010-65076599
地址: 北京市朝阳区劲松中街 3 号 100025

www.cttc.net.cn

检验报告
TEST REPORT

报告编号(No.): GS20000434-1 第 2 页共 6 页 Page 2of6

检验结论
Conclusion

| | | |
|----|--|------------|
| 1 | 基本要求 and 外观要求 General requirement and appearance | 符合 Pass |
| 2 | 耐干摩擦色牢度 Colour fastness to dry rubbing | — |
| 3 | 甲醛含量 Formaldehyde content | 符合 Pass |
| 4 | pH 值 pH value | 符合 Pass |
| 5 | 可分解致癌芳香胺染料 (24 种) Azo Dyes(24) | 符合 Pass |
| 6 | 可迁移性荧光增白物 Transferable fluorescent brighteners | 符合 Pass |
| 7 | 环氧乙烷残留量 Residual amount of ethylene oxide | 符合 Pass |
| 8 | 鼻夹长度 Nose clip length | 符合 Pass |
| 9 | 鼻夹耐折性 Nose clip folding resistance | 符合 Pass |
| 10 | 口罩带及口罩带与口罩体的连接处断裂强力 Breaking strength of the mask band and the joint of mask band and the mask body | 符合 Pass |
| 11 | 颗粒物过滤效率 Particle filtration efficiency | 符合 Pass |
| 12 | 细菌过滤效率 Bacterial filtration efficiency | 符合 Pass |
| 13 | 阻燃性能-燃烧时间 Flammability-burning time | 符合 Pass |
| 14 | 尖端和边缘锐利性 Sharpness of the point and edge | 符合 Pass |
| 15 | 微生物指标 Microorganism index | 符合 Pass |

检验结果详见附页
The conclusion please refer to the attached pages

中纺标检认证股份有限公司
中国商业联合会轻纺海品质量监督检测中心
进出口商品检验鉴定机构
国家纺织制品质量监督检验中心

Chinatesta Textile Testing & Certification Services Tel: 010-65002813 400-066-0486
Add: No.3 Yanjingli Middle Street, Chaoyang District, Beijing 10025, China Fax: 010-65076599
地址: 北京市朝阳区劲松中街 3 号 100025

www.cttc.net.cn

一次性使用医用口罩 (EN 14683 Type I)
Single-use Medical Face Mask(EN 14683 Type I)

Test Report No.: 721654570
Report Date: 18 May 2020

SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Anhui Tiangkang Medical Technology Co.,Ltd

CLIENT ADDRESS No.228 WeiyiRoad Economic Development Zone Tianchang City 239300
Anhui China

TEST PERIOD 01-May-2020~12-May-2020

Prepared By

(Bella Xu)
Report Drafter

Authorized By

(Lior Liu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested.(3) The test report shall not be reproduced except in full without the written approval of the laboratory.(4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TUV SUD Page 1 of 8

Test Report No.: 721654570
Report Date: 18 May 2020

TEST REPORT

Sample Description : Single-use Medical Face Mask
Sample Quantity : 50 pieces
Lot Number/Batch Code : 200416
Specification : With earloop
Size : /
Brand Name : /
Remark: The above information was provided by applicant.

Summary of Test Results

| No. | Test Item | Test Method | Test Standard Type I | Judgement |
|-----|--|-------------------------------------|----------------------|-----------|
| 1 | Bacterial Filtration Efficiency Test (BFE), % | EN 14683:2019+AC:2019(E) Annex B | ≥ 95 | Pass |
| 2 | Differential Pressure Test (Pa/cm ²) | EN 14683:2019+AC:2019(E) Annex C | < 40 | Pass |
| 3 | Microbial Cleanliness Test (CFU/g) | EN 14683:2019+AC:2019(E) Annex D | ≤ 30 | Pass |

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.


Photo of Samples

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108
P.R. China


Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TUV SUD Page 2 of 8



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01554-02A

Sample Name: Single-use medical face mask

Study Title: Skin Irritation Test - 0.9% Sodium Chloride
Injection Extract

Standard: ISO 10993-10:2010

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.

C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiankang Medical Technology Co.,Ltd

NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jssmt.com

Page 1 of 11

Report No.: SSMT-R-2020-01554-02A

1.0 Purpose

New Zealand rabbits were used to evaluate the potential of skin irritation of samples under the condition of this test.

2.0 Reference

Biological evaluation of medical devices-Part 10:Tests for irritation and skin sensitization(ISO 10993-10:2010)

Biological evaluation of medical devices-Part 12:Sample preparation and reference materials(ISO 10993-12:2012)

Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask

Sterilization state: Sterilized, EO

Model: Type I

Size: 17.5×9.5 cm

Lot/ Batch#: 200415

Physical State: Solid

Color: See the photo

Density: N/S

Stability: N/S

Solubility: N/S

Test Article Material: N/S


Packing Material: N/S

Storage Condition: Room temperature


Manufacturer: Anhui Tiankang Medical Technology Co.,Ltd.

Manufacturer address: NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China


Sample photograph:



Page 7 of 11



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01554-03A

Sample Name: Single-use medical face mask

Study Title: Skin Sensitization Test - 0.9% Sodium Chloride
Injection Extract

Standard: ISO 10993-10:2010

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.

C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiankang Medical Technology Co.,Ltd

NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: 86-519-83587899 www.jssmt.com

Page 1 of 12

Report No.: SSMT-R-2020-01554-03A

1.0 Purpose

The test was designed to evaluate the potential of a test article to cause skin sensitization using Guinea Pig Maximization Test.

2.0 Reference

Biological evaluation of medical devices-Part 10:Tests for irritation and skin sensitization(ISO 10993-10:2010)

Biological evaluation of medical devices-Part 12:Sample preparation and reference materials(ISO 10993-12:2012)

Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask

Sterilization state: Sterilized, EO

Model: Type I

Size: 17.5×9.5 cm

Lot/ Batch#: 200415

Physical State: Solid

Color: See the photo

Density: N/S

Stability: N/S

Solubility: N/S

Test Article Material: N/S


Packing Material: N/S

Storage Condition: Room temperature


Manufacturer: Anhui Tiankang Medical Technology Co.,Ltd.

Manufacturer address: NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China


Sample photograph:



Page 7 of 12



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01554-01A

Sample Name: single-use medical face mask

Study Title: In Vitro Cytotoxicity Test

Standard: ISO 10993-5:2009

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiankang Medical Technology Co., Ltd
NO. 288, WeiyiRoad, EconomicDevelopment Zone, Tianchang City, Anhui Province, P.R. China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jssmt.com

Page 1 of 11

Report No.:SSMT-R-2020-01554-01A

1.0 Purpose

The purpose of the test is to determine the potential cytotoxicity toxicity of a mammalian cell culture (mouse fibroblast L-929 cells) in response to the test article.

2.0 Standard


Biological evaluation of medical devices Part 5: Tests for In Vitro Cytotoxicity (ISO 10993-5:2009)
Biological evaluation of medical devices Part 12: Sample preparation and reference materials (ISO 10993-12:2012)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: single-use medical face mask
Sterilization state: Sterilized, EO
Model: Type I
Size: 17.5×9.5 cm
Lot/ Batch#: 200415
Physical State: Solid
Color: See the photo
Density: N/S
Stability: N/S
Solubility: N/S
Test Article Material: N/S
Packing Material: N/S
Storage Condition: Room temperature
Manufacturers: Anhui Tiankang Medical Technology Co., Ltd
Manufacturer address: NO. 288, WeiyiRoad, EconomicDevelopment Zone, Tianchang City, Anhui Province, P.R. China

Sample photograph:



3.2 Control Articles

3.2.1 Negative Control Article Name: High Density Polyethylene
Manufacturer: Jiangsu haiaosihui biotechnology co., LTD.
Size: 1.6 mm thick, 300*300 mm

Page 7 of 11

一次性使用医用口罩 (EN 14683 Type II)
Single-use Medical Face Mask(EN 14683 Type II)

Test Report No.: 721655016
Report Date: 29 May 2020

SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Anhui Tiangkang Medical Technology Co.,Ltd

CLIENT ADDRESS No.228 Weyi Road Development Zone Tianchang City 239300 Anhui China

TEST PERIOD 17-May-2020~25-May-2020

Prepared By

Bella Xu

(Bella Xu)
Report Drafter

Authorized By

(Leo Liu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TUV SUD Page 1 of 8

Test Report No.: 721655016
Report Date: 29 May 2020

TEST REPORT

Sample Description : Single-use Medical Face Mask
Sample Quantity : 50 pieces
Lot Number/Batch Code : 200428
Specification : /
Size : /
Brand Name : /
Remark: The above information was provided by applicant.

Summary of Test Results

| No. | Test Item | Test Method | Test Standard Type II | Judgement |
|-----|--|-------------------------------------|-----------------------|-----------|
| 1 | Bacterial Filtration Efficiency Test (BFE), % | EN 14683:2019+AC:2019(E) Annex B | ≥ 98 | Pass |
| 2 | Differential Pressure Test (Pa/cm ²) | EN 14683:2019+AC:2019(E) Annex C | < 40 | Pass |
| 3 | Microbial Cleanliness Test (CFU/g) | EN 14683:2019+AC:2019(E) Annex D | ≤ 30 | Pass |

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.


Photo of Samples

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China


Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TUV SUD Page 2 of 8



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01553-02A

Sample Name: Single-use medical face mask

Study Title: Skin Irritation Test - 0.9% Sodium Chloride
Injection Extract

Standard: ISO 10993-10:2010

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.

C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiankang Medical Technology Co.,Ltd.

NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jssmt.com

Page 1 of 11

Report No.: SSMT-R-2020-01553-02A

1.0 Purpose

New Zealand rabbits were used to evaluate the potential of skin irritation of samples under the condition of this test.

2.0 Reference

Biological evaluation of medical devices-Part 10:Tests for irritation and skin sensitization(ISO 10993-10:2010)

Biological evaluation of medical devices-Part 12:Sample preparation and reference materials(ISO 10993-12:2012)

Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask

Sterilization state: Sterilized, EO

Model: Type 11

Size: 17.5×9.5cm

Lot/ Batch#: 200415

Physical State: Solid

Color: See the photo

Density: N/S

Stability: N/S

Solubility: N/S

Test Article Material: N/S


Packing Material: N/S

Storage Condition: Room temperature


Manufacturer: Anhui Tiankang Medical Technology Co.,Ltd.

Manufacturer address: NO.288, WeiyiRoad, Economic Development Zone, Tianchang City, Anhui Province, P.R.China



Sample photograph:



Page 7 of 11



180015344189

中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01553-03A

Sample Name: Single-use medical face mask

Study Title: Skin Sensitization Test - 0.9% Sodium Chloride
Injection Extract

Standard: ISO 10993-10:2010

Test facility

Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiankang Medical Technology Co.,Ltd.
NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jssmt.com

Page 1 of 11

Report No.: SSMT-R-2020-01553-03A

1.0 Purpose

The test was designed to evaluate the potential of a test article to cause skin sensitization using Guinea Pig Maximization Test.

2.0 Reference


Biological evaluation of medical devices-Part 10:Tests for irritation and skin sensitization(ISO 10993-10:2010)
Biological evaluation of medical devices-Part 12:Sample preparation and reference materials(ISO 10993-12:2012)
Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles


3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask
Sterilization state: Sterilized, EO
Model: Type I I
Size: 17.5×9.5cm
Lot/ Batch#: 200415
Physical State: Solid
Color: See the photo
Density: N/S
Stability: N/S
Solubility: N/S
Test Article Material: N/S
Packing Material: N/S
Storage Condition: Room temperature
Manufacturer: Anhui Tiankang Medical Technology Co.,Ltd.
Manufacturer address: NO.288, WeiyiRoad, Economic Development Zone, Tianchang City, Anhui Province, P.R.China


Sample photograph:



Page 7 of 11



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-02310-01A

Sample Name: Single-use medical face mask

Study Title: In Vitro Cytotoxicity Test

Standard: ISO 10993-5:2009

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiakang Medical Technology Co., Ltd
NO. 288, WeiyiRoad, EconomicDevelopment Zone, Tianchang City, Anhui Province, P.R.China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jssmt.com

Page 1 of 11

Report No.:SSMT-R-2020-02310-01A

1.0 Purpose

The purpose of the test is to determine the potential cytotoxicity toxicity of a mammalian cell culture (mouse fibroblast L-929 cells) in response to the test article.

2.0 Standard


Biological evaluation of medical devices Part 5: Tests for In Vitro Cytotoxicity (ISO 10993-5:2009)
Biological evaluation of medical devices Part 12: Sample preparation and reference materials (ISO 10993-12:2012)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask
Sterilization state: Sterilized, EO
Model: Type I1
Size: 17.5*9.5
Lot/ Batch#: 200415
Physical State: Solid
Color: See the photo
Density: N/S
Stability: N/S
Solubility: N/S
Test Article Material: N/S
Packing Material: N/S
Storage Condition: Room temperature
Manufacturers: Anhui Tiakang Medical Technology Co., Ltd
Manufacturer address: NO. 288, WeiyiRoad, EconomicDevelopment Zone, Tianchang City, Anhui Province, P.R.China

Sample photograph:



3.2 Control Articles

3.2.1 Negative Control Article Name: High Density Polyethylene
Manufacturer: Jiangsu haiaosihui biotechnology co., LTD.
Size: 1.6 mm thick, 300*300 mm

Page 7 of 11

一次性使用医用口罩 (EN 14683 Type IIR)
Single-use Medical Face Mask(EN 14683 Type IIR)

Test Report No.: 721653819
Report Date: 27 April 2020

SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Anhui Tiangkang Medical Technology Co.,Ltd

CLIENT ADDRESS No.228 Weiyi Road Development Zone Tianchang City 239300 Anhui China

TEST PERIOD 10-Apr -2020~18-Apr-2020

Prepared By

(Bella Xu)
Report Drafter

Authorized By

(Leo Xu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food_chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

Page 1 of 11

Test Report No.: 721653819
Report Date: 27 April 2020

TEST REPORT

Sample Description : Single-use Medical Face Mask
Sample Quantity : 45 pieces
Lot Number/Batch Code : 200308
Specification : With earloop
Size : /
Type of Mask : Type IIR
Brand Name : /
Remark: The above information was provided by applicant.

Summary of Test Results

| No. | Test Item | Test Standard | Judgement |
|-----|--|----------------------------------|-----------|
| 1 | Bacterial Filtration Efficiency (BFE) Test | EN 14683:2019+AC:2019(E) Annex B | Pass |
| 2 | Differential Pressure Test | EN 14683:2019+AC:2019(E) Annex C | Pass |
| 3 | Synthetic Blood Penetration Test | ISO 22609:2004 | Pass |
| 4 | Microbial Cleanliness Test | EN 14683:2019+AC:2019(E) Annex D | Pass |

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.



Photo of Samples

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food_chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

Page 2 of 11

中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01555-02A

Sample Name: Single-use medical face mask

Study Title: Skin Irritation Test - 0.9% Sodium Chloride
Injection Extract

Standard: ISO 10993-10:2010

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.

C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiangkang Medical Technology Co.,Ltd.

NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China. 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jsssmi.com

Page 1 of 11

Report No.: SSMT-R-2020-01555-02A

1.0 Purpose

New Zealand rabbits were used to evaluate the potential of skin irritation of samples under the condition of this test.

2.0 Reference

Biological evaluation of medical devices-Part 10:Tests for irritation and skin sensitization(ISO 10993-10:2010)

Biological evaluation of medical devices-Part 12:Sample preparation and reference materials(ISO 10993-12:2012)

Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask

Sterilization state: Sterilized, EO

Model: Type IIR

Size: 17.5×9.5 cm

Lot/ Batch#: 200415

Physical State: Solid

Color: See the photo

Density: N/S

Stability: N/S

Solubility: N/S

Test Article Material: N/S


Packing Material: N/S

Storage Condition: Room temperature


Manufacturer: Anhui Tiangkang Medical Technology Co.,Ltd.

Manufacturer address: NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China


Sample photograph:



Page 7 of 11



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01555-03A

Sample Name: Single-use medical face mask

Study Title: Skin Sensitization Test - 0.9% Sodium Chloride
Injection Extract

Standard: ISO 10993-10:2010

| Test facility | Sponsor |
|---|---|
| Jiangsu Science Standard Medical Testing Co., Ltd. C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China | Anhui Tiakang Medical Technology Co.,Ltd. NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China |

Jiangsu Science Standard Medical Testing Co., Ltd.

C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: (86-519-83587899) www.jssmt.com

Page 1 of 12

Report No.: SSMT-R-2020-01555-03A

1.0 Purpose

The test was designed to evaluate the potential of a test article to cause skin sensitization using Guinea Pig Maximization Test.

2.0 Reference


Biological evaluation of medical devices-Part 10:Tests for irritation and skin sensitization(ISO 10993-10:2010)
 Biological evaluation of medical devices-Part 12:Sample preparation and reference materials(ISO 10993-12:2012)
 Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles


3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: Single-use medical face mask
 Sterilization state: Sterilized, EO
 Model: Type IIR
 Size: 17.5×9.5 cm
 Lot/ Batch#: 200415
 Physical State: Solid
 Color: See the photo
 Density: N/S
 Stability: N/S
 Solubility: N/S
 Test Article Material: N/S
 Packing Material: N/S
 Storage Condition: Room temperature
 Manufacturer: Anhui Tiakang Medical Technology Co.,Ltd.
 Manufacturer address: NO.288, Weiyi Road, Economic Development Zone, Tianchang City, Anhui Province, P.R.China


Sample photograph:



Page 7 of 12



180015344189



中国认可
国际互认
检测
TESTING
CNAS L10066

Test Report

Report Number: SSMT-R-2020-01555-01A

Sample Name: single-use medical face mask

Study Title: In Vitro Cytotoxicity Test

Standard: ISO 10993-5:2009

Test facility

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China

Sponsor

Anhui Tiankang Medical Technology Co., Ltd
NO. 288, WeiyiRoad, EconomicDevelopment Zone, Tianchang City, Anhui Province, P.R. China

Jiangsu Science Standard Medical Testing Co., Ltd.
C4 Building, No.9 Changyang Road, Wujin District, Changzhou, Jiangsu, China 213161 Tel: (86-519-83587899) Fax: 86-519-83587899 www.jssmt.com

Page 1 of 11

Report No.:SSMT-R-2020-01555-01A

1.0 Purpose

The purpose of the test is to determine the potential cytotoxicity toxicity of a mammalian cell culture (mouse fibroblast L-929 cells) in response to the test article.

2.0 Standard


Biological evaluation of medical devices Part 5: Tests for In Vitro Cytotoxicity (ISO 10993-5:2009)
Biological evaluation of medical devices Part 12: Sample preparation and reference materials (ISO 10993-12:2012)

3.0 Test and control articles

3.1 Test article (The information about the test article was supplied by the sponsor wherever applicable.)

Test article name: single-use medical face mask
Sterilization state: Sterilized, EO
Model: Type IIR
Size: 17.5×9.5 cm
Lot/ Batch#: 200415
Physical State: Solid
Color: See the photo
Density: N/S
Stability: N/S
Solubility: N/S
Test Article Material: N/S
Packing Material: N/S
Storage Condition: Room temperature
Manufacturers: Anhui Tiankang Medical Technology Co., Ltd
Manufacturer address: NO. 288, WeiyiRoad, EconomicDevelopment Zone, Tianchang City, Anhui Province, P.R. China

Sample photograph:



3.2 Control Articles

3.2.1 Negative Control Article Name: High Density Polyethylene
Manufacturer: Jiangsu haiaosihui biotechnology co., LTD.
Size: 1.6 mm thick, 300*300 mm

Page 7 of 11

自吸过滤式防颗粒物呼吸器
Non-powered Air-purifying Particle Respirator

171200110769

中国认可
国际互认
检测
TESTING
CNAS L0606

检 验 报 告

TEST REPORT

No: (2020) 皖检 XF 字 第 06230 号

产品名称
Product Name

自吸过滤式防颗粒物呼吸器

受检单位
Inspected Body

/

委托单位
Client

安徽天康医疗科技股份有限公司

检验类别
Kind of Test

委托检验

发
行
章

安徽省产品质量监督检验研究院
Anhui Provincial Supervising & Testing Research Institute for Product Quality

安徽省产品质量监督检验研究院

Anhui Provincial Supervising & Testing Research Institute for Product Quality

检 验 报 告

TEST REPORT

No: (2020) 皖检 XF 字 第 06230 号 共 3 页 第 1 页

| | | | | |
|-------------------------------------|-------------------------------|---|--|------------|
| 产品名称 Product Name | 自吸过滤式防颗粒物呼吸器 | 型号规格 Model/Type | / | |
| 标称生产单位 Nominal Manufacturer | / | 受检单位 Inspected Body | / | |
| 委托单位 Client | 安徽天康医疗科技股份有限公司 | 抽样单位 Sampling Body | / | |
| 委托单位地址 Client Address | / | 抽样地点 Sampling Site | / | |
| 检验项目 Test Items | 共查项【详见附页】 | 样品特性和状态 Sample Character and Condition | 外观无异常 | |
| 检验日期 Test Date | 2020.05.07-2020.05.10 | 原编号或生产日期 Serial Number/Manufactured Date | / | |
| 标称商标/品牌 Nominal Trade Mark/Brand | / | 抽样人员 Sampling Staff | 检查封样人员 Checking and Sealing Samples | / |
| 检验类别 Kind of Test | 委托检验 | 抽样基数 Sampling Base | 样品数量 Samples Quantity | 20 只 |
| 样品等级 Sample Grade | KN95 | 抽样日期 Sampling Date | 到样日期 Receipt Date | 2020.05.06 |
| 检验依据 Test Criteria | GB2626-2006 | | | |
| 检验结论 Test Conclusion | 该样品按 GB2626-2006 标准检验，所检项目合格。 | | | |
| 备注 Note | 样品信息和外观照片见附页。 | | | |

签发日期: 2020年7月1日

批准: 丁振华

Approved by:

审核: 王仁

Audited by:

主检: 纪扬扬

Tested by:

自吸过滤式防颗粒物呼吸器
Non-powered Air-purifying Particle Respirator

160010260248

(2016)国认监认字(099)号

中国认可
国际互认
检测
TESTING
CNAS L1499

检 验 报 告

检验报告编号: No. 2020(H) - 0487

样品名称: KN95 口罩

委托单位: 安徽天康医疗科技股份有限公司

检验类别: 委 托

国家劳动保护用品质量监督检验中心(北京)
(北京)

国家劳动保护用品质量监督检验中心(北京)
检 验 报 告

检验报告编号: No. 2020(H) - 0487 共 2 页 第 1 页

| | | | |
|---------|---|--------|----------------------------|
| 样品名称 | KN95 口罩 | 商 标 | _____ |
| 规格型号 | _____ | 防护级别 | _____ |
| 委托单位名称 | 安徽天康医疗科技股份有限公司 | 委托单位电话 | 0550-7309086 |
| 委托单位地址 | 安徽省天长市经济开发区纬一路 228 号 | | |
| 生产单位名称 | 安徽天康医疗科技股份有限公司 | 生产单位电话 | 0550-7308592 |
| 生产单位地址 | 安徽省天长市经济开发区纬一路 228 号 | | |
| 样品数量 | 20 | 生产日期 | _____ |
| 样品状态 | 完好 | 到样日期 | 2020.05.08 |
| 样品特征 | 白色折叠口罩 | 送样者 | 王东 |
| 检验依据 | GB 2626-2006《呼吸防护用品 自吸过滤式防颗粒物呼吸器》 | | |
| 检验项目 | 过滤效率、吸气阻力、呼气阻力 | | |
| 检 验 结 论 | <p>该样品经检验,依据 GB 2626-2006《呼吸防护用品 自吸过滤式防颗粒物呼吸器》,判定所检项目数据符合(KN95)标准要求。</p> <div style="text-align: right; margin-top: 10px;"> <p>(检验报告专用章) 签发日期: 2020 年 5 月 16 日</p> </div> | | |
| 备 注 | 样品未做预处理 | | |
| 批 准: | <u>陈辉为</u> | 审 核: | <u>冯 强</u> 主 检: <u>周芸芸</u> |

颗粒过滤半面罩 (EN 149 FFP3 NR)
Particle Filtering Half Mask (EN 149 FFP3 NR)



中国认可
国际互认
检测
TESTING
CNAS L7901

检验检测报告

TEST REPORT



STFWT202017468

Product Name Particle Filtering Half Mask

Trust Unit Anhui Tiangkang Medical Technology Co., Ltd.

Manufacturer Anhui Tiangkang Medical Technology Co., Ltd.

Test Category Entrusted Inspection




江苏省特种安全防护产品质量监督检验中心
JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS

Test Report

STFWT202017468 page 1 of 12

| | | | |
|--------------------------------|--|-----------------------|--------------|
| Product Name | Particle Filtering Half Mask | Specification Type | TK-5 |
| | | Trademark | TKMD |
| Trust Unit | Anhui Tiangkang Medical Technology Co., Ltd. | Tel. | 0550-7309156 |
| Manufacturer | Anhui Tiangkang Medical Technology Co., Ltd. | Sample Grade | FFP3 |
| Sample Quantity | 72 pcs | Sample Receiving Date | 2020-08-03 |
| Test Category | Entrusted inspection | Batch No./Article No | 20200710 |
| Samples Conditions | Meet the testing requirements | | |
| Document and Decide Accordance | EN 149: 2001+A1: 2009 《Respiratory protective devices -Filtering half masks to protect against particles-Requirements, testing, marking》 | | |
| Test Conclusion | The samples were tested, the items tested meet the requirements of EN 149:2001+A1:2009 standard for FFP3 level. Signature Date: 2020-08-28 | | |
| Remarks | The head harness of the mask provided by the applicant is head wearing. Compatibility with skin is not recognized by the center. The test data are only for reference. The sample is not marked for reuse (NR) and does not require testing for blocking performance. The test conclusion of this report is only for the items inspected and does not mean that the uninspected items or functions meet the requirements. The results apply to the sample as received. | | |

Approver 顾海燕 Examiner 杨森 Major tester 丁欣



JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS.

颗粒过滤半面罩带呼气阀 (EN 149 FFP3 NR)
Particle Filtering Half Mask With Exhalation Valve (EN 149 FFP3 NR)



中国认可
国际互认
检测
TESTING
CNAS L7901

检验检测报告

TEST REPORT



STFWT202017469

Product Name Particle Filtering Half Mask

Trust Unit Anhui Tiangkang Medical Technology Co., Ltd.

Manufacturer Anhui Tiangkang Medical Technology Co., Ltd.

Test Category Entrusted Inspection




江苏省特种安全防护产品质量监督检验中心
JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS

Test Report

STFWT202017469 page 1 of 13


| | | | |
|--------------------------------|--|-----------------------|--------------|
| Product Name | Particle Filtering Half Mask | Specification Type | TK-5V |
| | | Trademark | TKMD |
| Trust Unit | Anhui Tiangkang Medical Technology Co., Ltd. | Tel. | 0550-7309156 |
| Manufacturer | Anhui Tiangkang Medical Technology Co., Ltd. | Sample Grade | FFP3 |
| Sample Quantity | 72 pcs | Sample Receiving Date | 2020-08-03 |
| Test Category | Entrusted inspection | Batch No./Article No | 20200710 |
| Samples Conditions | Meet the testing requirements | | |
| Document and Decide Accordance | EN 149: 2001+A1: 2009 《Respiratory protective devices -Filtering half masks to protect against particles-Requirements, testing, marking》 | | |
| Test Conclusion | The samples were tested, the items tested meet the requirements of EN 149:2001+A1:2009 standard for FFP3 level. <div style="text-align: right;">Signature Date: 2020-08-28</div> | | |
| Remarks | The head harness of the mask provided by the applicant is head wearing. Compatibility with skin is not recognized by the center. The test data are only for reference. The sample is not marked for reuse (NR) and does not require testing for blocking performance. The test conclusion of this report is only for the items inspected and does not mean that the uninspected items or functions meet the requirements. The results apply to the sample as received. | | |

Approver 顾海燕 Examiner 杨森 Major tester 丁欣




JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS.

一次性使用医用口罩 (成人款)
Single-use Medical Face Mask(Adult)



170015143957




中国认可
国际互认
检测
TESTING
CNAS L9908

检 验 报 告

报告编号: AH2020-QSJ-00398

检 品 名 称: 一次性使用医用口罩

检 验 目 的: 委托检验



安徽省食品药品检验研究院

ZLJL-165-04

安徽省食品药品检验研究院 检验报告首页

报告编号: AH2020-QSJ-00398
共 3 页 第 1 页

| | | | |
|-------|---|---------|-----------------------|
| 样品名称 | 一次性使用医用口罩 | 样品编号 | AH2020-QSJ-00398 |
| | 送样 (√) 抽样 () | | |
| 商 标 | Tkmd | 型号规格 | 17.5cm x 9.5cm |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 委托检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号/批号 | 批200319 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200325 |
| 抽样单位 | / | 样品数量 | 100个 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/05/06 | 检验日期 | 2020/05/06~2020/05/26 |
| 检验项目 | 4.1-4.7.1 | | |
| 检验依据 | YY/T 0969-2013《一次性使用医用口罩》 | | |
| 检验结论 | 所检项目符合YY/T 0969-2013《一次性使用医用口罩》的要求 (检验报告专用章或检验单位公章) 签发日期: 2020/5/26 | | |
| 备 注 | 报告中“—”表示不适用项,“/”表示空白项。 | | |

批 准: 

职 务: 授权签字人

审 核: 

主 检: 



一次性使用医用口罩（儿童款）
Single-use Medical Face Mask(Children)

170015143957

中国认可
国际互认
检测
TESTING
CNAS L9908

检 验 报 告

报告编号: AH2020-QSJ-00397

检 品 名 称: 一次性使用医用口罩

检 验 目 的: 委托检验

安徽省食品药品检验研究院
检验报告专用章

ZLJL-165-04

安徽省食品药品检验研究院 检验报告首页

报告编号: AH2020-QSJ-00397
共 3 页 第 1 页

| | | | |
|-------|--|---------|-----------------------|
| 样品名称 | 一次性使用医用口罩 | 样品编号 | AH2020-QSJ-00397 |
| | 送样 (√) 抽样 () | | |
| 商 标 | Tkmd | 型号规格 | 儿童款 14.5cm x 9cm |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 委托检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号/批号 | 批200412 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200412 |
| 抽样单位 | / | 样品数量 | 100个 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/05/06 | 检验日期 | 2020/05/06~2020/05/26 |
| 检验项目 | 4.1-4.7.1 | | |
| 检验依据 | YY/T 0969-2013《一次性使用医用口罩》 | | |
| 检验结论 | 所检项目符合YY/T 0969-2013《一次性使用医用口罩》的要求。 (检验报告专用章或检验单位公章) 签发日期: 2020/5/26 | | |
| 备注 | 报告中“—”表示不适用项,“/”表示空白项。 | | |

批 准:


审 核:

主 检:

职 务: 授权签字人

医用外科口罩（非灭菌型 / 系带式）
Surgical Mask(Non-sterile/with Tie Coverall)

MA CNAS 中国认可 国际互认 检测 TESTING CNAS L9908 170015143957



扫码验证报告


检 验 报 告

报告编号: AH2020-QSJ-00496

检品名称: 医用外科口罩

检验目的: 委托检验

安徽省食品药品检验研究院



ZLJL-165-04



安徽省食品药品检验研究院 检验报告照片页


报告编号: AH2020-QSJ-00496 共3页第3页

| |
|---|
| 照片和说明 |
|  |
| 样品描述 |
| 型号规格或其它说明 |
| 18cm×9.5cm (±5%) |



医用外科口罩 (灭菌型 / 系带式)
Surgical Mask (Sterile / With Tie Coverall)





 中国认可
国际互认
检测
TESTING
CNAS L9908
170015143957


扫码验证报告

检 验 报 告

报告编号: AH2020-QSJ-00497


检品名称: 医用外科口罩
 检验目的: 委托检验



 安徽省食品药品检验研究院


ZLJL-165-04


安徽省食品药品检验研究院 检验报告照片页

报告编号: AH2020-QSJ-00497 共 3 页 第 3 页

| |
|--|
| 照片和说明 |
|  |
| 样品描述 |
| 型号规格或其它说明 |
| 18cm×9.5cm (±5%) |



一次性使用医用口罩 (符合欧盟 REACH 法令测试)
Single-use Medical Face Mask (Compliance with the EU REACH testing)



Test Report

Report No. : WP-20057115-JC-01En Page No. : 1 / 16

Client Name: Anhui Ti ankang Medical Technology Co., Ltd
Client Address: No.228 Weiyi Road Economic Development Zone Tianchang City

The following sample(s) was/were submitted and identified on behalf of the applicant as:

Date of Sample Received: 2020-05-11
Testing Period: 2020-05-11~2020-05-18
Test Requested: Selected test (s) as requested by client.
Test Criterion: Please refer to next page(s).
Test Results: Please refer to next page(s).


Summary: PASS According to the ruling of the Court of Justice of the European Union on the definition of an article under REACH, and the specified scope and evaluation screening, the test results of SVHC are ≤ 0.1% (w/w) in the articles of the submitted sample.

Complied by: 杨治宏

Inspected by: 强艳妮


Approved by: 商婉艳

Issued Date: 2020-05-18



Shanghai Micro-spectral Chemistry Analysis and Test technology Co.,Ltd.

Address: Building 9, Building 10, Building 18, Urban Industrial Park, Lane 139, Guowei Road, Yangpu District, Shanghai
 Tel: 400-776-7627 Web: www.weipujishu.com



Test Report

Report No. : WP-20057115-JC-01En Page No. : 2 / 16

Test Part Description

| No. | Sample Name | Sample ID | Description | Material/Brand |
|-----|------------------------------|-------------|--|----------------|
| 001 | Single-use Medical Face Mask | 200504474-1 | Silver metal | / |
| 002 | Single-use Medical Face Mask | 200504474-1 | Blue Face Mask (Non metal mixed test) | / |

I. Test Items: Total Cadmium (Cd) and Total Lead (Pb)
Test Method and Apparatus:

| Test Items | Test methods | Apparatus |
|--------------------|-----------------------------|-----------|
| Total Cadmium (Cd) | Laboratory internal methods | ICP-OES |
| Total Lead (Pb) | Laboratory internal methods | ICP-OES |

Test Results:


| Test Items | Unit | MDL | 001+002 |
|--------------------|-------|-----|---------|
| Total Lead(Pb) | mg/kg | 10 | N.D. |
| Total Cadmium (Cd) | mg/kg | 10 | N.D. |

Remarks:

- (1) 1mg/kg = 1ppm = 0.0001%
- (2) MDL =Method Detection Limit
- (3) N.D.=Not Detected (<MDL)
- (4) " - " =Not Regulated
- (5) When the result is equal to or greater than 500 mg/kg, the limit shall not apply where it can be demonstrated that the rate of Lead release from such an article or any such accessible part of an article, whether coated or uncoated, does not exceed 0.05 µg/cm²/h (equivalent to 0.05 µg/g/h). However, the legal text does not define a specific method for the demonstration of release rate while the Commission may develop a method for Lead migration.
- (6) The composite sampling method is based on the client's special request and the result is calculated using the minimum sample weight.

Address: Building 9, Building 10, Building 18, Urban Industrial Park, Lane 139, Guowei Road, Yangpu District, Shanghai
 Tel: 400-776-7627 Web: www.weipujishu.com

自吸过滤式防颗粒物呼吸器（符合欧盟 REACH 法令测试）
Non-powered Air-purifying Particle Respirator(Compliance with the EU REACH testing)



Test Report

Report No. : WP-20057115-JC-02En Page No. : 1 / 16

Client Name: Anhui Ti ankang Medical Technology Co., Ltd
Client Address: No.228 Weiyi Road Economic Development Zone Tianchang City

The following sample(s) was/were submitted and identified on behalf of the applicant as:

Date of Sample Received: 2020-05-11
Testing Period: 2020-05-11~2020-05-18
Test Requested: Selected test (s) as requested by client.
Test Criterion: Please refer to next page(s).
Test Results: Please refer to next page(s).


Summary: PASS According to the ruling of the Court of Justice of the European Union on the definition of an article under REACH, and the specified scope and evaluation screening, the test results of SVHC are ≤ 0.1% (w/w) in the articles of the submitted sample.

Complied by: 杨治宏

Inspected by: 张艳妮


Approved by: 肖婉艳

2020-05-18



Shanghai Micro-spectral Chemistry Analysis and Test technology Co.,Ltd.

Address: Building 9, Building 10, Building 18, Urban Industrial Park, Lane 139, Guowei Road, Yangpu District, Shanghai
 Tel: 400-776-7627 Web: www.weipujishu.com



Test Report

Report No. : WP-20057115-JC-02En Page No. : 2 / 16

Test Part Description

| No. | Sample Name | Sample ID | Description | Material/Brand |
|-----|----------------------------|-------------|---|----------------|
| 001 | Disposable Protective Mask | 200504476-1 | Silver metal | / |
| 002 | Disposable Protective Mask | 200504476-1 | White Face Mask (Non metal mixed test) | / |

I. Test Items: Total Cadmium (Cd) and Total Lead (Pb)
Test Method and Apparatus:

| Test Items | Test methods | Apparatus |
|--------------------|-----------------------------|-----------|
| Total Cadmium (Cd) | Laboratory internal methods | ICP-OES |
| Total Lead (Pb) | Laboratory internal methods | ICP-OES |

Test Results:

| Test Items | Unit | MDL | 001+002 |
|--------------------|-------|-----|---------|
| Total Lead(Pb) | mg/kg | 10 | N.D. |
| Total Cadmium (Cd) | mg/kg | 10 | N.D. |

Remarks:

- (1) 1mg/kg = 1ppm = 0.0001%
- (2) MDL =Method Detection Limit
- (3) N.D.=Not Detected (<MDL)
- (4) " - " =Not Regulated
- (5) When the result is equal to or greater than 500 mg/kg, the limit shall not apply where it can be demonstrated that the rate of Lead release from such an article or any such accessible part of an article, whether coated or uncoated, does not exceed 0.05 µg/cm²/h (equivalent to 0.05 µg/g/h). However, the legal text does not define a specific method for the demonstration of release rate while the Commission may develop a method for Lead migration.
- (6) The composite sampling method is based on the client's special request and the result is calculated using the minimum sample weight.

Address: Building 9, Building 10, Building 18, Urban Industrial Park, Lane 139, Guowei Road, Yangpu District, Shanghai
 Tel: 400-776-7627 Web: www.weipujishu.com

红外线额温计
Infrared Forehead Thermometer

检 验 报 告

报告编号: AH2020-QZC-00735

检品名称: 红外线额温计

检验目的: 注册检验

安徽省食品药品检验研究院



ZLJL-165-03

安徽省食品药品检验研究院 检验报告首页

报告编号: AH2020-QZC-00735

共55页 第1页

| | | | | |
|-------|---|---------|--------------------------|------------------|
| 样品名称 | 红外线额温计 | | 样品编号 | AH2020-QZC-00735 |
| | 送样 (√) | 抽样 () | | |
| 商 标 | Tkmd | 型号规格 | TKWQ-01 | |
| 委托方 | 安徽天康医疗科技股份有限公司 | | 检验类型 | 注册检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号/批号 | TKWQ012004010011/批200401 | |
| 生产单位 | 安徽天康医疗科技股份有限公司 | | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | | 生产日期 | 20200401 |
| 抽样单位 | / | 样品数量 | 1台 | |
| 抽样地点 | / | 抽样基数 | / | |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 | |
| 收样日期 | 2020/04/23 | 检验日期 | 2020/04/23~2020/05/15 | |
| 检验项目 | 全项目 | | | |
| 检验依据 | 安徽天康医疗科技股份有限公司《红外线额温计》 | | | |
| 检验结论 | 被检样品符合安徽天康医疗科技股份有限公司《红外线额温计》的要求 (检验报告专用章或检验报告专用章) 签发日期: 2020/5/18 | | | |
| 备 注 | 报告中“—”表示不适用,“/”表示空白,“*”表示分包项。本院不具备分包项检验能力,经委托方同意分包给分包方: 1. 国家家用电器产品质量监督检验中心(安徽),其资质认定许可编号为: 180008113609。 | | | |

批 准: 赵永昕 审 核: 孙谷玄 主 检: 杨盼
职 务: 授权签字人

医用防护口罩（灭菌型 / 折叠型）

Protective Face Mask for Medical Use (Sterile / Folded)



扫码验证报告

检验报告

报告编号: AH2020-QZC-00873

检品名称: 医用防护口罩

检验目的: 注册检验

安徽省食品药品检验研究院



ZLJL-165-04

安徽省食品药品检验研究院 检验报告首页

报告编号: AH2020-QZC-00873


共 3 页 第 1 页

| | | | |
|-------|--|-----------|-----------------------|
| 样品名称 | 医用防护口罩 | 样品编号 | AH2020-QZC-00873 |
| | 送样 (√) 抽样 () | | |
| 商 标 | Tkmd | 型号规格 | 灭菌版 |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 注册检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号 / 批号 | 200401 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200401 |
| 抽样单位 | / | 样品数量 | 120 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/05/25 | 检验日期 | 2020/05/25~2020/06/15 |
| 检验项目 | 2.1-2.7, 2.8.2-2.11 | | |
| 检验依据 | 安徽天康医疗科技股份有限公司《医用防护口罩》产品技术要求 | | |
| 检验结论 | 被检样品符合安徽天康医疗科技股份有限公司《医用防护口罩》产品技术要求。 (检验报告专用章或检验单位公章) 签发日期: 2020/6/15 | | |
| 备注 | 报告中“—”表示不适用项, “/”表示空白项。 | | |



批 准: 审 核: 主 检:
职 务: 授权签字人

医用防护口罩（非灭菌型 / 折叠型）
Protective Face Mask for Medical Use (Non-sterile / Folded)




扫码验证报告

检 验 报 告

报 告 编 号：AH2020-QZC-00874

检 品 名 称：医用防护口罩

检 验 目 的：注册检验



安徽省食品药品检验研究院
检验报告专用章

ZLJL-165-04

安徽省食品药品检验研究院 检 验 报 告 首 页

报告编号：AH2020-QZC-00874
共 3 页 第 1 页

| | | | |
|-------|--|-----------|-----------------------|
| 样品名称 | 医用防护口罩 | 样品编号 | AH2020-QZC-00874 |
| | 送样 (√) 抽样 () | | |
| 商 标 | Tkmd | 型号规格 | 非灭菌型 折叠型 |
| 委托方 | 安徽天康医疗科技股份有限公司 | 检验类别 | 注册检验 |
| 委托方地址 | 安徽省天长市经济开发区纬一路228号 | 产品编号 / 批号 | 200401 |
| 生产单位 | 安徽天康医疗科技股份有限公司 | 抽样单编号 | / |
| 受检单位 | 安徽天康医疗科技股份有限公司 | 生产日期 | 20200401 |
| 抽样单位 | / | 样品数量 | 50 |
| 抽样地点 | / | 抽样基数 | / |
| 抽样日期 | / | 检验地点 | 安徽省食品药品检验研究院 |
| 收样日期 | 2020/05/25 | 检验日期 | 2020/05/25~2020/06/08 |
| 检验项目 | 2.8.1 | | |
| 检验依据 | 安徽天康医疗科技股份有限公司《医用防护口罩》产品技术要求 | | |
| 检验结论 | 所检项目符合安徽天康医疗科技股份有限公司《医用防护口罩》产品技术要求。 (检验报告专用章或检验单位公章) 签发日期：2020/6/8 | | |
| 备 注 | 报告中“—”表示不适用项，“/”表示空白项。 | | |

批 准： 

职 务： 授权签字人

审 核： 

主 检： 



安徽省食品药品检验研究院
检验报告专用章

一次性使用医用口罩 (非灭菌型 / 松紧带式)
Single-use Medical Face Mask(Non-sterile/With Earloop)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率≥ 95%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 可展开褶皱处理，获得更优防护

可选颜色：



Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|------------------------|------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 初包装 First package | 19x10x8.5cm/50pcs |
| | 大包装 Outer package | 53x40x36.5cm/2000pcs |

一次性使用医用口罩 (非灭菌型 / 系带式)
Single-use Medical Face Mask(Non-sterile/With Tie Coverall)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率≥ 95%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|--------------------------|--------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 330x170mm/25pcs |
| | 中包装 Secondary package | 21x12.5x13cm/50pcs |
| | 大包装 Outer package | 65.5x43.5x27.5cm/1000pcs |

一次性使用医用口罩 (灭菌型 / 松紧带式)
Single-use Medical Face Mask (Sterile/With Earloop)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率≥ 95%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|------------------------------------|--------------------------|------------------------|
| 灭菌型 / 松紧带式 Sterile/With earloop | 小包装 Unit package | 218x103mm |
| | 中包装 Secondary package | 21.8x10.5x20cm/50pcs |
| | 大包装 Outer package | 56x45.5x41.5cm/1000pcs |

一次性使用医用口罩 (灭菌型 / 系带式)
Single-use Medical Face Mask (Sterile/With Tie Coverall)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率≥ 95%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|--|--------------------------|------------------------|
| 灭菌型 / 系带式 Sterile/With tie coverall | 小包装 Unit package | 218x103mm/1pc |
| | 中包装 Secondary package | 21.8x10.5x11cm/20pcs |
| | 大包装 Outer package | 55.5x45x35cm/600pcs |

一次性使用医用口罩 (灭菌型 / 儿童款)
Single-use Medical Face Mask (Sterile/Children)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率≥ 95%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|-------------------------------|--------------------------|--------------------------|
| 灭菌型 / 儿童款 Sterile/Children | 小包装 Unit package | 185x103mm |
| | 中包装 Secondary package | 18.5x10.5x20cm/50pcs |
| | 大包装 Outer package | 55.5x38.5x41.5cm/1000pcs |

一次性使用医用口罩 (非灭菌型 / 儿童款)
Single-use Medical Face Mask (Non-sterile/Children)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率≥ 95%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|------------------------------------|------------------------|------------------------|
| 非灭菌型 / 儿童款 Non-sterile/Children | 初包装 First package | 16x10x8.5cm/50pcs |
| | 大包装 Outer package | 53x33.5x36.5cm/2000pcs |

医用外科口罩 (非灭菌型 / 耳挂式)
Surgical Mask(Non-sterile/With Earloop)



适用范围:

供临床医务人员在有创操作过程中佩带, 覆盖住使用者的口、鼻及下颌, 为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能:

- 细菌过滤效率≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation ,covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE≥95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|--|------------------------|------------------------|
| 非灭菌型 / 耳挂式 Non-sterile/With earloop | 初包装 First package | 19x10x8.5cm/50pcs |
| | 大包装 Outer package | 53x40x36.5cm/2000pcs |

医用外科口罩 (非灭菌型 / 系带式)
Surgical Mask(Non-sterile/With Tie Coverall)



适用范围:

供临床医务人员在有创操作过程中佩带, 覆盖住使用者的口、鼻及下颌, 为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能:

- 细菌过滤效率≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation ,covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE≥95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|--------------------------|--------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 330x170mm/25pcs |
| | 中包装 Secondary package | 21x12.5x13cm/50pcs |
| | 大包装 Outer package | 65.5x43.5x27.5cm/1000pcs |

医用外科口罩 (灭菌型 / 耳挂式)
Surgical Mask (Sterile / With Earloop)



适用范围：

供临床医务人员在有创操作过程中佩带，覆盖住使用者的口、鼻及下颌，为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能：

- 细菌过滤效率 ≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation, covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE ≥ 95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|-------------------------------------|--------------------------|------------------------|
| 灭菌型 / 耳挂式 Sterile / With earloop | 小包装 Unit package | 218x103mm |
| | 中包装 Secondary package | 21.8x10.5x20cm/50pcs |
| | 大包装 Outer package | 56x45.5x41.5cm/1000pcs |

医用外科口罩 (灭菌型 / 系带式)
Surgical Mask (Sterile / With Tie Coverall)



适用范围：

供临床医务人员在有创操作过程中佩带，覆盖住使用者的口、鼻及下颌，为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能：

- 细菌过滤效率 ≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹，密合性好
- 三层防护，呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理，获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation, covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE ≥ 95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|--|--------------------------|------------------------|
| 灭菌型 / 系带式 Sterile / With tie coverall | 小包装 Unit package | 218x103mm/1pc |
| | 中包装 Secondary package | 21.8x10.5x11cm/20pcs |
| | 大包装 Outer package | 55.5x45x35cm/600pcs |

一次性使用医用口罩 (EN 14683 Type I, 非灭菌型 / 松紧带式)
 Single-use Medical Face Mask(EN 14683 Type I,Non-sterile/With earloop)



适用范围 :

适用于患者和其他人在流行病或传染病的情况下佩戴。覆盖使用者的口、鼻和下颌,阻隔从口腔和鼻腔呼出或喷出的污染物,以降低感染传播的风险。

产品性能 :

- 细菌过滤效率≥ 95%
- 透气性 (压力差) < 40Pa/cm²
- 微生物清洁度≤ 30cfu/g
- 可调式鼻夹, 密性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for patients and other persons to wear it in epidemic or pandemic situations,covering the user's mouth, nose and jaw, blocking the exhaling or spraying pollutants from the mouth and nasal cavity to reduce the risk of infection transmission.

Product Performance:

- BFE≥95%.
- Differential Pressure<40Pa/cm².
- Microbial cleanliness≤30cfu/g.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|------------------------|------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 初包装 First package | 19x10x8.5cm/50pcs |
| | 大包装 Outer package | 53x40x36.5cm/2000pcs |

一次性使用医用口罩 (EN 14683 Type I, 非灭菌型 / 系带式)
 Single-use Medical Face Mask(EN 14683 Type I,Non-sterile/With tie coverall)



适用范围 :

适用于患者和其他人在流行病或传染病的情况下佩戴。覆盖使用者的口、鼻子和下巴,阻隔从口腔和鼻腔呼出或喷洒的污染物,以降低感染传播的风险。

产品性能 :

- 细菌过滤效率≥ 95%
- 透气性 (压力差) < 40Pa/cm²
- 微生物清洁度≤ 30cfu/g
- 可调式鼻夹, 密性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for patients and other persons to wear it in epidemic or pandemic situations,covering the user's mouth, nose and jaw, blocking the exhaling or spraying pollutants from the mouth and nasal cavity to reduce the risk of infection transmission.

Product Performance:

- BFE≥95%.
- Differential Pressure<40Pa/cm².
- Microbial cleanliness≤30cfu/g.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|--------------------------|--------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 330x170mm/25pcs |
| | 中包装 Secondary package | 21x12.5x13cm/50pcs |
| | 大包装 Outer package | 65.5x43.5x27.5cm/1000pcs |

一次性使用医用口罩 (EN 14683 Type II, 非灭菌型 / 松紧带式)
Single-use Medical Face Mask(EN 14683 Type II,Non-sterile/With earloop)



适用范围:

适用于在外科手术过程中和其他具有类似要求的医疗环境中的工作人员使用,以限制工作人员和患者之间传染病的传播。

产品性能:

- 细菌过滤效率≥ 98%
- 透气性(压力差) < 40Pa/cm²
- 微生物清洁度≤ 30cfu/g
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

The product intended to limit the transmission of infective agents between staff and patients during surgical procedures and other medical settings with similar requirements.

Product Performance:

- BFE≥98%.
- Differential Pressure<40Pa/cm².
- Microbial cleanliness≤30cfu/g.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|------------------------|------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 初包装 First package | 19x10x8.5cm/50pcs |
| | 大包装 Outer package | 53x40x36.5cm/2000pcs |

一次性使用医用口罩 (EN 14683 Type II, 非灭菌型 / 系带式)
Single-use Medical Face Mask(EN 14683 Type II,Non-sterile/With tie coverall)



适用范围:

适用于在外科手术过程中和其他具有类似要求的医疗环境中的工作人员使用,以限制工作人员和患者之间传染病的传播。

产品性能:

- 细菌过滤效率≥ 98%
- 透气性(压力差) < 40Pa/cm²
- 微生物清洁度≤ 30cfu/g
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

The product intended to limit the transmission of infective agents between staff and patients during surgical procedures and other medical settings with similar requirements.

Product Performance:

- BFE≥98%.
- Differential Pressure<40Pa/cm².
- Microbial cleanliness≤30cfu/g.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|--------------------------|--------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 330x170mm/25pcs |
| | 中包装 Secondary package | 21x12.5x13cm/50pcs |
| | 大包装 Outer package | 65.5x43.5x27.5cm/1000pcs |

一次性使用医用口罩 (EN 14683 Type IIR, 非灭菌型 / 耳挂式)
 Single-use Medical Face Mask(EN 14683 Type IIR,Non-sterile/With earloop)



适用范围:

适用于在外科手术过程中和其他具有类似要求的医疗环境中的工作人员使用,以限制工作人员和患者之间传染病的传播。

产品性能:

- 细菌过滤效率≥ 98%
- 透气性 (压力差) < 60Pa/cm²
- 微生物清洁度≤ 30cfu/g
- 抗飞溅性 (抗合成血液穿透) ≥ 16.0kPa
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

The product intended to limit the transmission of infective agents between staff and patients during surgical procedures and other medical settings with similar requirements.

Product Performance:

- BFE≥98%.
- Differential Pressure<60Pa/cm².
- Microbial cleanliness≤30cfu/g.
- Splash resistance pressure≥16.0kPa.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|------------------------|------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 初包装 First package | 19x10x8.5cm/50pcs |
| | 大包装 Outer package | 53x40x36.5cm/2000pcs |

一次性使用医用口罩 (EN 14683 Type IIR, 非灭菌型 / 系带式)
 Single-use Medical Face Mask(EN 14683 Type IIR,Non-sterile/With tie coverall)



适用范围:

适用于在外科手术过程中和其他具有类似要求的医疗环境中的工作人员使用,以限制工作人员和患者之间传染病的传播。

产品性能:

- 细菌过滤效率≥ 98%
- 透气性 (压力差) < 60Pa/cm²
- 微生物清洁度≤ 30cfu/g
- 抗飞溅性 (抗合成血液穿透) ≥ 16.0kPa
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:


The product intended to limit the transmission of infective agents between staff and patients during surgical procedures and other medical settings with similar requirements.

Product Performance:

- BFE≥98%.
- Differential Pressure<60Pa/cm².
- Microbial cleanliness≤30cfu/g.
- Splash resistance pressure≥16.0kPa.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---|--------------------------|--------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 330x170mm/25pcs |
| | 中包装 Secondary package | 21x12.5x13cm/50pcs |
| | 大包装 Outer package | 65.5x43.5x27.5cm/1000pcs |

一次性使用医用口罩 (非灭菌型 / 松紧带式, 纸塑小袋)
 Single-use Medical Face Mask(Non-Sterile/With Earloop, Blister bag)



适用范围:

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能:

- 细菌过滤效率≥ 95%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 纸塑小袋尺寸 Blister bag Size |
|---|----------------------------------|----------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 1 只 / 纸塑小袋 1pc/Blister bag | 240x130mm |
| | 10 只 / 纸塑小袋 10pcs/Blister bag | 250x145mm |
| | 20 只 / 纸塑小袋 20pcs/Blister bag | 300x170mm |
| | 50 只 / 纸塑小袋 50pcs/Blister bag | 430x350mm |

一次性使用医用口罩 (灭菌型 / 松紧带式, 纸塑小袋)
 Single-use Medical Face Mask(Sterile/With Earloop, Blister bag)



适用范围:

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能:

- 细菌过滤效率≥ 95%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 纸塑小袋尺寸 Blister bag Size |
|------------------------------------|----------------------------------|----------------------------|
| 灭菌型 / 松紧带式 Sterile/With earloop | 1 只 / 纸塑小袋 1pc/Blister bag | 240x130mm |
| | 10 只 / 纸塑小袋 10pcs/Blister bag | 250x145mm |
| | 20 只 / 纸塑小袋 20pcs/Blister bag | 300x170mm |
| | 50 只 / 纸塑小袋 50pcs/Blister bag | 430x350mm |

医用外科口罩 (灭菌型 / 耳挂式, 纸塑小袋)
Surgical Mask (Sterile/With Earloop, Blister bag)



适用范围:

供临床医务人员在有创操作过程中佩戴, 覆盖住使用者的口、鼻及下颌, 为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能:

- 细菌过滤效率 ≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation, covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE ≥ 95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 纸塑小袋尺寸 Blister bag Size |
|-----------------------------------|----------------------------------|----------------------------|
| 灭菌型 / 耳挂式 Sterile/With earloop | 1 只 / 纸塑小袋 1pc/Blister bag | 240x130mm |
| | 10 只 / 纸塑小袋 10pcs/Blister bag | 250x145mm |
| | 20 只 / 纸塑小袋 20pcs/Blister bag | 300x170mm |
| | 50 只 / 纸塑小袋 50pcs/Blister bag | 430x350mm |

医用外科口罩 (非灭菌型 / 耳挂式, 纸塑小袋)
Surgical Mask (Non-Sterile/With Earloop, Blister bag)



适用范围:

供临床医务人员在有创操作过程中佩戴, 覆盖住使用者的口、鼻及下颌, 为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能:

- 细菌过滤效率 ≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation, covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE ≥ 95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 纸塑小袋尺寸 Blister bag Size |
|--|----------------------------------|----------------------------|
| 非灭菌型 / 耳挂式 Non-sterile/With earloop | 1 只 / 纸塑小袋 1pc/Blister bag | 240x130mm |
| | 10 只 / 纸塑小袋 10pcs/Blister bag | 250x145mm |
| | 20 只 / 纸塑小袋 20pcs/Blister bag | 300x170mm |
| | 50 只 / 纸塑小袋 50pcs/Blister bag | 430x350mm |

一次性使用医用口罩 (非灭菌型 / 松紧带式, OPP 复合膜带悬挂孔小袋)
 Single-use Medical Face Mask(Non-Sterile/With Earloop, polybag with hanging hole)



适用范围:

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能:

- 细菌过滤效率≥ 95%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:


It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | OPP 复合膜小袋尺寸 Polybag Size |
|---|--------------------------|-----------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 小包装 Unit package | 280x150mm/10pcs |
| | 中包装 Secondary package | 280x205x150mm/100pcs |
| | 大包装 Outer package | 575x425x470mm/1200pcs |

一次性使用医用口罩 (非灭菌型 / 系带式, OPP 复合膜带悬挂孔小袋)
 Single-use Medical Face Mask(Non-Sterile/with tie coverall, polybag with hanging hole)



适用范围:

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能:

- 细菌过滤效率≥ 95%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE≥95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | OPP 复合膜小袋尺寸 Polybag Size |
|---|--------------------------|-----------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 295x150mm/10pcs |
| | 中包装 Secondary package | 29.5x18x15cm/60pcs |
| | 大包装 Outer package | 60.5x37.5x47cm/720pcs |

医用外科口罩 (非灭菌型 / 系带式, OPP 复合膜带悬挂孔小袋)
Surgical Mask (Non-Sterile/with tie coverall, polybag with hanging hole)



适用范围 :

供临床医务人员在有创操作过程中佩戴, 覆盖住使用者的口、鼻及下颌, 为防止病原体、微生物、体液、颗粒物等的直接透过提供物理屏障。

产品性能 :

- 细菌过滤效率 ≥ 95%
- 非油性颗粒过滤效率不小于 30%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 具有抗合成血液穿透性
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for clinical medical staff to wear it during invasive operation, covering the user's mouth, nose and jaw, to provide a physical barrier to prevent direct transmission of pathogens, microorganisms, body fluids, particulate matter, etc.

Product Performance:

- BFE ≥ 95%.
- The PFE is not less than 30%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- It has resistance against penetration by synthetic blood.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | OPP 复合膜小袋尺寸 Polybag Size |
|---|--------------------------|-----------------------------|
| 非灭菌型 / 系带式 Non-sterile/With tie coverall | 小包装 Unit package | 295x150mm/10pcs |
| | 中包装 Secondary package | 29.5x18x15cm/60pcs |
| | 大包装 Outer package | 60.5x37.5x47cm/720pcs |

一次性使用医用口罩 (EN 14683 Type I, 非灭菌型 / 松紧带式, OPP 复合膜带悬挂孔小袋)

Single-use Medical Face Mask (EN 14683 Type I, Non-Sterile/With Earloop, polybag with hanging hole)



适用范围 :

适用于患者和其他人在流行病或传染病的情况下佩戴。覆盖使用者的口、鼻和下颌, 阻隔从口腔和鼻腔呼出或喷出的污染物, 以降低感染传播的风险。

产品性能 :

- 细菌过滤效率 ≥ 95%
- 透气性 (压力差) < 40Pa/cm²
- 微生物清洁度 ≤ 30cfu/g
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for patients and other persons to wear it in epidemic or pandemic situations, covering the user's mouth, nose and jaw, blocking the exhaling or spraying pollutants from the mouth and nasal cavity to reduce the risk of infection transmission.

Product Performance:

- BFE ≥ 95%.
- Differential Pressure < 40Pa/cm².
- Microbial cleanliness ≤ 30cfu/g.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | OPP 复合膜小袋尺寸 Polybag Size |
|---|--------------------------|-----------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 小包装 Unit package | 280x150mm/10pcs |
| | 中包装 Secondary package | 28x20.5x15cm/100pcs |
| | 大包装 Outer package | 57.5x42.5x47cm/1200pcs |

一次性使用医用口罩 (EN 14683 Type IIR, 非灭菌型 / 耳挂式, OPP 复合膜带悬挂孔小袋)

Single-use Medical Face Mask (EN 14683 Type IIR, Non-Sterile/With Earloop, polybag with hanging hole)



适用范围：

适用于在外科手术过程中和其他具有类似要求的医疗环境中的工作人员使用,以限制工作人员和患者之间传染病的传播。

产品性能：

- 细菌过滤效率 ≥ 98%
- 透气性 (压力差) < 60Pa/cm²
- 微生物清洁度 ≤ 30cfu/g
- 抗喷溅性 (抗合成血液穿透) ≥ 16.0kPa
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

The product intended to limit the transmission of infective agents between staff and patients during surgical procedures and other medical settings with similar requirements.

Product Performance:

- BFE ≥ 98%.
- Differential Pressure < 60Pa/cm².
- Microbial cleanliness ≤ 30cfu/g.
- Splash resistance pressure ≥ 16.0kPa.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | OPP 复合膜小袋尺寸 Polybag Size |
|---|--------------------------|-----------------------------|
| 非灭菌型 / 松紧带式 Non-sterile/With earloop | 小包装 Unit package | 280x150mm/10pcs |
| | 中包装 Secondary package | 28x20.5x15cm/100pcs |
| | 大包装 Outer package | 57.5x42.5x47cm/1200pcs |

一次性使用医用口罩 (EN 14683 Type I, 儿童款, 非灭菌型 / 松紧带式, OPP 复合膜带悬挂孔小袋)

Single-use Medical Face Mask (EN 14683 Type I, Children, Non-sterile/With Earloop, Polybag With Hanging Hole)



适用范围：

用于佩戴者在不存在体液和喷溅风险的普通医疗环境下的卫生护理。

产品性能：

- 细菌过滤效率 ≥ 95%
- 可调式鼻夹, 密合性好
- 三层防护, 呼吸顺畅
- 可展开褶皱处理, 获得更优防护

Intended Use:

It is used for the health care of the wearer in a general medical environment where there is no risk of body fluids and splashing.

Product Performance:

- BFE ≥ 95%.
- Adjustable nose clip, good fit.
- Three layers of protection, smooth breathing.
- Expandable fold processing for better protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 纸塑小袋尺寸 Blister bag Size |
|------------------------------------|----------------------------------|----------------------------|
| 非灭菌型 / 儿童款 Non-sterile/Children | 1 只 / 纸塑小袋 1pc/Blister bag | 200x120mm |
| | 10 只 / 纸塑小袋 10pcs/Blister bag | 240x150mm |
| | 20 只 / 纸塑小袋 20pcs/Blister bag | 270x170mm |
| | 50 只 / 纸塑小袋 50pcs/Blister bag | 320x260mm |

红外线额温计

Infrared Forehead Thermometer



适用范围：

通过测量额头的热辐射来显示被测对象的体温。

产品性能：

- 非接触式测量，避免交叉感染
- 进口传感器
- 三色背光体温警示，方便夜间测量，体温警示更直接
- 双 32 组记忆，掌握体温连续变化，支持临床热型诊断
- 发烧报警，一机多测

Intended Use:

To display the body temperature by measuring the thermal radiation on the forehead.

Product Performance:

- Non-contact measurement to avoid cross infection.
- Inlet sensor.
- Three-color backlight temperature warning, convenient night measurement, temperature warning more direct.
- Double 32 group memory, grasp the continuous changes of body temperature, support clinical thermal diagnosis.
- Fever alarm, one machine with multiple functions.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|----------------------|------------------------|------------------------|
| TKWQ-01 | 初包装 First package | 100x50x160mm/1pc |
| | 大包装 Outer package | 33.5x32x28cm/30pcs |

医用防护口罩（灭菌型 / 折叠型）

Protective Face Mask for Medical use (Sterile/Folded)



适用范围：

用于戴在医疗机构与病毒物料接触的人员面部，用于防止来自患者的病毒向医务人员传播。

产品性能：

- 非油性颗粒过滤效率等级 1 级 (≥ 95%)
- 具有抗合成血液穿透性
- 可调式鼻夹，密合性好
- 立体透气，畅快呼吸

Intended Use:

Used to wear on the face of persons in contact with virus materials in medical institutions, to prevent the transmission of virus from patients to medical personnel.

Product Performance:

- Non-oily particle filtration efficiency Class 1 (≥95%).
- It has resistance against penetration by synthetic blood.
- Adjustable nose clip, good fit.
- 3D ventilation, smooth breathing.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|-----------------------------|--------------------------|-------------------------|
| 灭菌型 / 折叠型 Sterile/Folded | 小包装 Unit package | 210x120mm /1pc |
| | 中包装 Secondary package | 21x12x14cm/25pcs |
| | 大包装 Outer package | 49.5x43.5x43.5cm/600pcs |

医用防护口罩（非灭菌型 / 折叠型）

Protective Face Mask for Medical use(Non-Sterile/Folded)



适用范围：

用于戴在医疗机构与病毒物料接触的人员面部，用于防止来自患者的病毒向医务人员传播。

产品性能：

- 非油性颗粒过滤效率等级 1 级 (≥ 95%)
- 具有抗合成血液穿透性
- 可调式鼻夹，密合性好
- 立体透气，畅快呼吸

Intended Use:

Used to wear on the face of persons in contact with virus materials in medical institutions, to prevent the transmission of virus from patients to medical personnel.

Product Performance:

- Non-oily particle filtration efficiency Class 1 (≥95%).
- It has resistance against penetration by synthetic blood.
- Adjustable nose clip , good fit.
- 3D ventilation, smooth breathing.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|----------------------------------|--------------------------|------------------------|
| 非灭菌型 / 折叠型 Non-Sterile/Folded | 小包装 Unit package | 175x125mm /1pc |
| | 中包装 Secondary package | 17.5x12.5x14cm/25pcs |
| | 大包装 Outer package | 54x39x29.5cm/450pcs |

自吸过滤式防颗粒物呼吸器（TK-3, 折叠型）

Non-powered Air-purifying Particle Respirator(TK-3,Folded)



适用范围：

适用于对空气中非油性颗粒物的呼吸防护，包括各种机械力或热产生的粉尘、烟和雾等。

产品性能：

- KN95 过滤布，非油性颗粒物过滤效率 ≥ 95%
- 五层防护，安全净化
- 立体裁切，全方位贴合保护
- 内置密封条，深层阻隔，呼吸顺畅
- 弹性耳带，佩戴舒适

Scope of Application:

Applicable to the respiratory protection of non-oily particulate matter in the air, including dust, smoke and fog generated by various mechanical forces or heat.

Product Performance:

- KN95 filter fabric, the filtration efficiency of non-oily particles ≥95%
- Five layers of protection and safety purification.
- 3D cutting, comprehensive fit protection
- With sponge strips for deep blocking, breathing smoothly.
- Elastic earloop for comfortable wearing.

| 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|--------------------------|--------------------------|
| 小包装 Unit package | 175x125m/1pc |
| 中包装 Secondary Package | 27x17.5x12.5cm/50pcs |
| 大包装 Outer package | 65.5x55.5x54.5cm/1500pcs |

颗粒过滤半面罩 (EN 149 FFP2 NR,TK-3 型)
Particle Filtering Half Mask (EN 149 FFP2 NR,TK-3)



适用范围：

适用于对某些颗粒物的呼吸防护，包括矿物质或煤尘、棉尘、面粉尘等，也适用于不会产生有害挥发性气体的液体或非油性颗粒。

产品性能：

- FFP2 级防护，用于某些油性和非油颗粒物的呼吸防护，过滤效率不低于 94%
- 五层防护，安全净化
- 立体裁切，全方位贴合保护
- 内置密封条，深层阻隔，呼吸顺畅
- 弹性耳带，佩戴舒适

Scope of Application:

It is used for respiratory protection for some particulates, such as mineral or coal dust, cotton dust, flour dust, etc. It is also suitable for liquid or non oily particles that do not produce harmful volatile gases.

Product Performance:

- The protection level is FFP2, it is used for respiratory protection of some oily and non oil particles, the penetration of the filter of the filtering half mask is not less than 94%.
- Five layers of protection and safety purification.
- 3D cutting, comprehensive fit protection
- With sponge strips for deep blocking, breathing smoothly
- Elastic earloop for comfortable wearing

| 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|--------------------------|------------------------|
| 小包装 Unit package | 205x130mm/1pc |
| 中包装 Secondary Package | 205x130x130cm/20pcs |
| 大包装 Outer package | 680x425x680cm/1000pcs |

医用隔离眼罩
Medical Isolation Goggles



适用范围：

适用于医疗机构中检查治疗时起防护作用，阻隔体液、血液飞溅或泼溅。

产品性能：

- 良好的透气性
- 镜片透光性好、防飞溅、防冲击
- 弹性头箍，可调节长度，适合各种头型佩戴
- 一体式镜片，可与矫视眼镜一同佩戴
- 舒适服帖，密封防护

Intended Use:

It is suitable for medical institutions to play a protective role during inspection and treatment, to blocking body fluids, blood splashing or splashing.

Product Performance:

- Good air permeability.
- The ocular has good light transmittance, splash proof and impact proof.
- Elastic head hoop, adjustable length, suitable for all kinds of heads.
- Integrated ocular, which can be worn together with the straightening spectacle.
- Comfortable fit, sealed protection.

| 产品类型 Product Type | 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|---------------------------|--------------------------|------------------------|
| 隔离眼罩 Isolation Goggles | 小包装 Unit package | 280x190mm |
| | 中包装 Secondary package | 10x7x19cm/1pc |
| | 大包装 Outer package | 53x45.5x39.5cm/60pcs |

医用一次性防护服
Medical Coveralls



适用范围：

适用于为医护人员在工作时接触具有潜在感染性的患者血液、体液、分泌物、空气中的颗粒物等提供阻隔、防护作用。

产品性能：

- 透湿量不小于 2500g/(m²·d)
- 抗合成血液穿透性不低于 1.75KPa
- 防护服关键部位材料及接缝处对非油性颗粒的过滤效率不小于 70%
- 带电量不大于 0.6μC/ 件

Intended Use:

Provide barriers and protection for medical personnel in contact with potentially infectious patients' blood, body fluids, secretions, and particulates in the air, etc.

Product Performance:

The moisture permeability is not less than 2500g/(m²·d).
The resistance against penetration by synthetic blood is not less than 1.75KPa.
The PFE which is in the key parts and seam of the non-woven coveralls is not less than 70%.
Electric capacity not more than 0.6 μ C/ piece.

| 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|------------------------|------------------------|
| 小包装 Unit package | 460x300mm |
| 大包装 Outer package | 51x37x38cm/20 件 |

隔离衣
Non-woven Isolation Gowns



适用范围：

适用于医疗机构门诊、病房、检验室等作普通隔离。

产品性能：

- 透气性能≥ 80mm/s
- 非油性颗粒过滤效率不小于 30%
- 沾水等级不低于 3 级

Intended Use:

Applicable to general isolation of medical institutions, clinics, wards, laboratory, etc.

Product Performance:

Air Permeability≥80mm/s.
The PFE is not less than 30%.
The grade of resistance to surface wetting is not lower than 3.

| 包装类型 Packaging Type | 包装尺寸 Packaging Size |
|------------------------|------------------------|
| 小包装 Unit package | 430x310mm |
| 大包装 Outer package | 54x37x45cm/40 件 |

医用口罩生产线 (1号车间)

Production Line of Medical Face Mask (No. 1 workshop)



医用口罩生产线 (2号车间)

Production Line of Medical Face Mask (No. 2 workshop)



防护服生产线

Production Line of Medical Coveralls





安徽天康(集团)股份有限公司

邮编(Zip): 239300

电话(Tel): +86-550-7777777

传真(Fax): +86-550-7028077

邮箱(Mail): xsc@tiankang.com

网址(Http): www.tiankang.com

地址: 安徽省天长市仁和南路20号

ADD:No.20 South Renhe Road, Tianchang City, Anhui
Province China

安徽天康医疗科技股份有限公司

邮编(Zip): 239300

电话(Tel): +86-550-7309165

传真(Fax): +86-550-7309158

邮箱(Mail): BBD@tkmedical.com

网址(Http): www.tkmedical.com

地址: 安徽省天长市经济开发区纬一路228号

ADD:No.228 Weiyi Road Economic Development Zone
Tianchang City Anhui China